

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD042921\  
 Data File : PD062554.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 29 Apr 2021 17:40  
 Operator : AR\AJ  
 Sample : M2161-04MS  
 Misc :  
 ALS Vial : 38 Sample Multiplier: 1

Instrument :  
 ECD\_D  
 ClientSampleId :  
 C0AQ3MS

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Apr 30 02:59:23 2021  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD042921CLP.M  
 Quant Title : GC Extractables  
 QLast Update : Fri Apr 30 02:49:51 2021  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1 µl  
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
-----						
System Monitoring Compounds						
1) SA Tetrachlo...	3.441	3.973	17778467	240.5E6	18.570	19.806
27) SA Decachlor...	8.203	9.036	37364813	377.9E6	34.810	34.621
Target Compounds						
2) A alpha-BHC	3.878	4.362	41916470	524.3E6	29.938	27.725
3) MA gamma-BHC...	4.161	4.647	39870321	503.0E6	29.847	29.115
4) MA Heptachlor	4.463	5.155	38970652	494.9E6	30.350	28.433
5) MB Aldrin	4.711	5.460	38671075	495.7E6	29.688	28.655
6) B beta-BHC	4.412	4.817	17863555	218.4E6	28.773	28.611
7) B delta-BHC	4.617	5.034	17263942	238.7E6	12.688	12.864
8) B Heptachlo...	5.154	5.846	37430889	455.5E6	30.544	28.400
9) A Endosulfan I	5.493	6.202	33363260	433.4E6	29.837	29.546
10) B trans-Chl...	5.382	6.080	37461091	459.7E6	29.881	28.066
11) B cis-Chlor...	5.440	6.152	36242002	448.1E6	29.514	28.870
12) B 4,4'-DDE	5.604	6.310	76124588	921.7E6	61.512	58.821
13) MA Dieldrin	5.737	6.459	79470857	944.8E6	63.046	60.931
14) MA Endrin	5.995	6.677	69990949	775.2E6	66.149	62.158
15) B Endosulfa...	6.269	6.887	67039035	800.5E6	64.640	63.965
16) A 4,4'-DDD	6.121	6.800	63676413	794.3E6	63.947	63.909
17) MA 4,4'-DDT	6.363	7.099	61602401	721.8E6	65.769	60.794
18) B Endrin al...	6.437	7.011	56844496	649.1E6	63.709	59.494
19) B Endosulfa...	6.651	7.234	54932949	676.3E6	54.954	52.999
20) A Methoxychlor	6.909	7.557	172.3E6	1720.7E6	322.936	303.580
21) B Endrin ke...	7.143	7.704	79803841	773.5E6	68.097	62.851

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD042921\  
 Data File : PD062554.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 29 Apr 2021 17:40  
 Operator : AR\AJ  
 Sample : M2161-04MS  
 Misc :  
 ALS Vial : 38 Sample Multiplier: 1

Instrument :  
 ECD\_D  
 ClientSampleId :  
 C0AQ3MS

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Apr 30 02:59:23 2021  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD042921CLP.M  
 Quant Title : GC Extractables  
 QLast Update : Fri Apr 30 02:49:51 2021  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1 µl  
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

