

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042921\
Data File : PD062578.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Apr 2021 23:15
Operator : AR\AJ
Sample : M2065-14
Misc :
ALS Vial : 59 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
BG567

Manual Integrations
APPROVED

Ankita
4/30/2021 2:31:45 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 30 03:05:44 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD042921CLP.M
Quant Title : GC Extractables
QLast Update : Fri Apr 30 02:49:51 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



QEdit

(16) 4,4'-DDD (A)
6.122min 1.279 ng/ml
response 1273937

(16) 4,4'-DDD #2 (A)
6.801min 0.037 ng/ml
response 462480

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042921\
Data File : PD062578.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Apr 2021 23:15
Operator : AR\AJ
Sample : M2065-14
Misc :
ALS Vial : 59 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
BG567

Manual Integrations
APPROVED

Ankita
4/30/2021 2:31:45 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 30 03:05:44 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD042921CLP.M
Quant Title : GC Extractables
QLast Update : Fri Apr 30 02:49:51 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(16) 4,4'-DDD (A)
6.121min 1.544 ng/ml m
response 1537866

(16) 4,4'-DDD #2 (A)
6.799min 3.005 ng/ml m
response 37350422

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042921\
 Data File : PD062578.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Apr 2021 23:15
 Operator : AR\AJ
 Sample : M2065-14
 Misc :
 ALS Vial : 59 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 BG567

Manual Integrations
 APPROVED

Ankita
 4/30/2021 2:31:45 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 30 03:05:44 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD042921CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Apr 30 02:49:51 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.441	3.973	12811691	154.4E6	13.382	12.718
27)	SA Decachlor...	8.204	9.037	25873693	257.6E6	24.105	23.604
Target Compounds							
12)	B 4,4'-DDE	5.604	6.310	2787162	58738234	2.252	3.749 #
16)	A 4,4'-DDD	6.121	6.799	1537866	37350422	1.544m	3.005m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

AT
 05/05/21