

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD043021\
 Data File : PD062636.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Apr 2021 16:25
 Operator : AR\AJ
 Sample : M2192-01
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 BG595

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 30 23:20:12 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD042921CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Apr 30 02:49:51 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.441	3.972	11555446	113.1E6	12.070	9.314m
27)	SA Decachlor...	8.203	9.034	28233380	209.5E6	26.303	19.190m#
Target Compounds							
5)	MB Aldrin	4.713	5.454	868320	23546659	0.667	1.361m#
8)	B Heptachlo...	5.152	5.855	1072652	43280924	0.875m	2.698 #
10)	B trans-Chl...	5.367	6.066	1309845	142.9E6	1.045m	8.725 #
12)	B 4,4'-DDE	5.621	6.303	1057901	14132590	0.855	0.902
15)	B Endosulfa...	6.269	6.887	2083856	27658884	2.009m	2.210m
16)	A 4,4'-DDD	6.120	6.792	2391547	71732291	2.402m	5.771m#
18)	B Endrin al...	6.450	7.006	1309201	10638050	1.467m	0.975 #
19)	B Endosulfa...	6.663	7.242	8849110	26555515	8.853m	2.081m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD043021\
Data File : PD062636.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Apr 2021 16:25
Operator : AR\AJ
Sample : M2192-01
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
BG595

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 30 23:20:12 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD042921CLP.M
Quant Title : GC Extractables
QLast Update : Fri Apr 30 02:49:51 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

