

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050323\
Data File : PD075187.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 May 2023 00:40
Operator : AR\AJ
Sample : 02447-06
Misc :
ALS Vial : 29 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
EW911

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/06/2023
Supervised By :Ankita Jodhani 05/08/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 04 03:00:17 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(3) gamma-BHC (Lindane) (MA)

4.344min -5.664 ng/ml

response -20362931

(3) gamma-BHC (Lindane) #2 (MA)

3.609min 4.368 ng/ml

response 6420362

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050323\
Data File : PD075187.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 May 2023 00:40
Operator : AR\AJ
Sample : 02447-06
Misc :
ALS Vial : 29 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
EW911

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/06/2023
Supervised By :Ankita Jodhani 05/08/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 04 03:00:17 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(3) gamma-BHC (Lindane) (MA)

4.342min 0.991 ng/ml m

response 3562117

(3) gamma-BHC (Lindane) #2 (MA)

3.609min 4.368 ng/ml

response 6420362

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050323\
Data File : PD075187.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 May 2023 00:40
Operator : AR\AJ
Sample : 02447-06
Misc :
ALS Vial : 29 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
EW911

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/06/2023
Supervised By :Ankita Jodhani 05/08/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 04 03:00:17 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(7) delta-BHC (B)

4.762min 0.151 ng/ml

response 511860

(7) delta-BHC #2 (B)

4.141min 9.090 ng/ml

response 14363806

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050323\
Data File : PD075187.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 May 2023 00:40
Operator : AR\AJ
Sample : 02447-06
Misc :
ALS Vial : 29 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
EW911

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/06/2023
Supervised By :Ankita Jodhani 05/08/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 04 03:00:17 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(7) delta-BHC (B)

4.801min 5.617 ng/ml m

response 18986938

(7) delta-BHC #2 (B)

4.141min 9.090 ng/ml

response 14363806

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050323\
Data File : PD075187.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 May 2023 00:40
Operator : AR\AJ
Sample : 02447-06
Misc :
ALS Vial : 29 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
EW911

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/06/2023
Supervised By :Ankita Jodhani 05/08/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 04 03:00:17 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(14) Endrin (MA)

6.591min 14.534 ng/ml

response 42029529

(14) Endrin #2 (MA)

5.646min 3.541 ng/ml

response 4307096

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050323\
Data File : PD075187.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 May 2023 00:40
Operator : AR\AJ
Sample : 02447-06
Misc :
ALS Vial : 29 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
EW911

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/06/2023
Supervised By :Ankita Jodhani 05/08/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 04 03:00:17 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(14) Endrin (MA)

6.591min 14.534 ng/ml

response 42029529

(14) Endrin #2 (MA)

5.607min 17.028 ng/ml m

response 20711582

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050323\
Data File : PD075187.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 May 2023 00:40
Operator : AR\AJ
Sample : 02447-06
Misc :
ALS Vial : 29 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
EW911

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/06/2023
Supervised By :Ankita Jodhani 05/08/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 04 03:00:17 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(17) 4,4'-DDT (MA)

7.043min 2.863 ng/ml

response 7268238

(17) 4,4'-DDT #2 (MA)

6.034min 24.829 ng/ml

response 25024596

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050323\
Data File : PD075187.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 May 2023 00:40
Operator : AR\AJ
Sample : 02447-06
Misc :
ALS Vial : 29 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
EW911

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/06/2023
Supervised By :Ankita Jodhani 05/08/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 04 03:00:17 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(17) 4,4'-DDT (MA)

7.014min 19.844 ng/ml m

response 50368161

(17) 4,4'-DDT #2 (MA)

6.032min 25.910 ng/ml m

response 26113436

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050323\
Data File : PD075187.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 May 2023 00:40
Operator : AR\AJ
Sample : 02447-06
Misc :
ALS Vial : 29 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
EW911

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/06/2023
Supervised By :Ankita Jodhani 05/08/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 04 03:00:17 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(27) Decachlorobiphenyl (SA)

9.093min 22.132 ng/ml

response 54770596

(27) Decachlorobiphenyl #2 (SA)

7.919min 21.244 ng/ml

response 20062290

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050323\
Data File : PD075187.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 May 2023 00:40
Operator : AR\AJ
Sample : 02447-06
Misc :
ALS Vial : 29 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
EW911

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/06/2023
Supervised By :Ankita Jodhani 05/08/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 04 03:00:17 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(27) Decachlorobiphenyl (SA)

9.092min 28.637 ng/ml m

response 70867491

(27) Decachlorobiphenyl #2 (SA)

7.919min 21.244 ng/ml

response 20062290