

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050323\
 Data File : PD075204.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 May 2023 08:56
 Operator : AR\AJ
 Sample : 02447-20RE
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 EW938RE

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 05/05/2023
 Supervised By : Ankita Jodhani 05/08/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 04 11:10:00 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu May 04 02:14:07 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.542	2.772	79182377	11463297	34.716m	10.868 #
27)	SA Decachlor...	9.085	7.915	163.7E6	51387086	66.132	54.413
Target Compounds							
7)	B delta-BHC	4.804	4.141	4784.3E6	1923.0E6	1415.410m	1216.985
14)	MA Endrin	6.592	5.607	5213.5E6	2010.1E6	1802.871	1652.631m
17)	MA 4,4'-DDT	7.015	6.033	5234.5E6	2109.4E6	2062.267m	2092.975
19)	B Endosulfa...	7.157	6.332	254.9E6	123.0E6	92.676m	105.786m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050323\
Data File : PD075204.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 May 2023 08:56
Operator : AR\AJ
Sample : 02447-20RE
Misc :
ALS Vial : 42 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 04 11:10:00 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Manual Integrations
APPROVED
Reviewed By: Abdul Mirza 05/05/2023
Supervised By: Ankita Jodhani 05/08/2023

Instrument :
ECD_D
ClientSampled :
EW/938RE