

Method Path : P:\HPCHEM1\ECD_D\Method\
 Method File : DTX050518.M
 Title : GC Extractables
 Last Update : Sat May 05 00:21:10 2018
 Response Via : Initial Calibration

Calibration Files

500 =PD047443.D 1000 =PD047441.D 750 =PD047442.D
 250 =PD047444.D =

	Compound	500	1000	750	250	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.123	1.102	1.130	1.122	1.119	E6	1.09
2)	Toxaphene-1	8.729	8.455	8.605	8.762	8.638	E3	1.61
3)	Toxaphene-2	2.248	2.152	2.240	2.246	2.221	E4	2.09
4)	Toxaphene-3	2.564	2.400	2.506	2.423	2.473	E4	3.06
5)	Toxaphene-4	2.275	2.116	2.181	2.235	2.202	E4	3.14
6)	Toxaphene-5	2.861	2.718	2.838	2.776	2.798	E4	2.29
7) SA	Decachlorobiphenyl	1.259	1.204	1.245	1.286	1.249	E6	2.75

Signal #2 Calibration Files

500 =PD047443.D 1000 =PD047441.D 750 =PD047442.D
 250 =PD047444.D =

	Compound	500	1000	750	250	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.775	1.662	1.747	1.861	1.761	E7	4.66
2)	Toxaphene-1	1.105	1.009	1.017	1.191	1.080	E5	7.94
3)	Toxaphene-2	1.922	1.705	1.859	1.970	1.864	E5	6.20
4)	Toxaphene-3	5.398	4.997	5.296	5.668	5.340	E5	5.19
5)	Toxaphene-4	3.328	3.208	3.354	3.657	3.387	E5	5.65
6)	Toxaphene-5	2.170	2.296	2.391	2.291	2.287	E5	3.94
7) SA	Decachlorobiphenyl	1.908	1.739	1.832	2.035	1.878	E7	6.66

(#) = Out of Range