

Data Path : P:\HPCHEM1\ECD D\Data\PD050518\
 Data File : PD047441.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 May 2018 20:28
 Operator : UA/AJ
 Sample : PTOXICC1000
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PTOXICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 05 00:29:32 2018
 Quant Method : P:\HPCHEM1\ECD_D\Method\DTX050518.M
 Quant Title : GC Extractables
 QLast Update : Sat May 05 00:17:19 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.428	4.041	110.2E6	1661.9E6	99.058	96.721
7) SA Decachlor...	8.148	9.120	120.4E6	1739.1E6	97.750	95.368
Target Compounds						
2) Toxaphene-1	5.662	6.431	8455407	100.9E6	984.082	954.338
3) Toxaphene-2	6.262	7.127	21521254	170.5E6	978.261	940.059
4) Toxaphene-3	6.507	7.202	23998518	499.7E6	966.940	961.385
5) Toxaphene-4	6.858	7.290	21155528	320.8E6	963.645	981.533
6) Toxaphene-5	6.991	7.697	27183801	229.6E6	974.499	1028.113

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : P:\HPCHEM1\ECD D\Data\PD050518\
Data File : PD047441.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 May 2018 20:28
Operator : UA/AJ
Sample : PTOXICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampled :
PTOXICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 05 00:29:32 2018
Quant Method : P:\HPCHEM1\ECD_D\Method\DTX050518.M
Quant Title : GC Extractables
QLast Update : Sat May 05 00:17:19 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Ini. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

