

Data Path : P:\HPCHEM1\ECD D\Data\PD050518\
 Data File : PD047443.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 May 2018 20:57
 Operator : UA/AJ
 Sample : PTOXICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PTOXICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 05 00:27:56 2018
 Quant Method : P:\HPCHEM1\ECD_D\Method\DTX050518.M
 Quant Title : GC Extractables
 QLast Update : Sat May 05 00:16:05 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.428	4.041	56136932	887.3E6	50.000	50.000
7) SA Decachlor...	8.147	9.119	62961006	954.0E6	50.000	50.000
Target Compounds						
2) Toxaphene-1	5.661	6.431	4364474	55261013	500.000	500.000
3) Toxaphene-2	6.261	7.126	11238883	96108957	500.000	500.000
4) Toxaphene-3	6.505	7.200	12819781	269.9E6	500.000	500.000
5) Toxaphene-4	6.857	7.289	11375888	166.4E6	500.000	500.000
6) Toxaphene-5	6.990	7.696	14303253	108.5E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : P:\HPCHEM1\ECD D\Data\PD050518\
Data File : PD047443.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 May 2018 20:57
Operator : UA/AJ
Sample : PTOXICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampled :
PTOXICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 05 00:27:56 2018
Quant Method : P:\HPCHEM1\ECD_D\Method\DTX050518.M
Quant Title : GC Extractables
QLast Update : Sat May 05 00:16:05 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Ini. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

