

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050523\
Data File : PD075307.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2023 02:58
Operator : AR\AJ
Sample : 02479-15
Misc :
ALS Vial : 45 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
EW974

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/09/2023
Supervised By :Ankita Jodhani 05/09/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 06 06:10:38 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(12) 4,4'-DDE (B)

6.205min 19.125 ng/ml

response 62455828

(12) 4,4'-DDE #2 (B)

5.234min 4.366 ng/ml

response 5612433

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050523\
Data File : PD075307.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2023 02:58
Operator : AR\AJ
Sample : 02479-15
Misc :
ALS Vial : 45 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
EW974

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/09/2023
Supervised By :Ankita Jodhani 05/09/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 06 06:10:38 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(12) 4,4'-DDE (B)

6.205min 19.125 ng/ml

response 62455828

(12) 4,4'-DDE #2 (B)

5.232min 10.698 ng/ml m

response 13753123

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050523\
Data File : PD075307.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2023 02:58
Operator : AR\AJ
Sample : 02479-15
Misc :
ALS Vial : 45 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
EW974

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/09/2023
Supervised By :Ankita Jodhani 05/09/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 06 06:10:38 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(13) Dieldrin (MA)

6.362min 16.262 ng/ml

response 55611742

(13) Dieldrin #2 (MA)

5.364min 48.033 ng/ml

response 65534898

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050523\
Data File : PD075307.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2023 02:58
Operator : AR\AJ
Sample : 02479-15
Misc :
ALS Vial : 45 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
EW974

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/09/2023
Supervised By :Ankita Jodhani 05/09/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 06 06:10:38 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(13) Dieldrin (MA)

6.361min 14.550 ng/ml m

response 49758339

(13) Dieldrin #2 (MA)

5.363min 94.369 ng/ml m

response 128755600

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050523\
Data File : PD075307.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2023 02:58
Operator : AR\AJ
Sample : 02479-15
Misc :
ALS Vial : 45 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
EW974

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/09/2023
Supervised By :Ankita Jodhani 05/09/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 06 06:10:38 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(14) Endrin (MA)

6.588min 44.041 ng/ml

response 127357257

(14) Endrin #2 (MA)

5.642min 11.960 ng/ml

response 14546623

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050523\
Data File : PD075307.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2023 02:58
Operator : AR\AJ
Sample : 02479-15
Misc :
ALS Vial : 45 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
EW974

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/09/2023
Supervised By :Ankita Jodhani 05/09/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 06 06:10:38 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(14) Endrin (MA)

6.587min 42.438 ng/ml m

response 122721613

(14) Endrin #2 (MA)

5.605min 45.715 ng/ml m

response 55603502

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050523\
Data File : PD075307.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2023 02:58
Operator : AR\AJ
Sample : 02479-15
Misc :
ALS Vial : 45 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
EW974

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/09/2023
Supervised By :Ankita Jodhani 05/09/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 06 06:10:38 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(17) 4,4'-DDT (MA)

7.040min 3.225 ng/ml

response 8185633

(17) 4,4'-DDT #2 (MA)

6.031min 86.441 ng/ml

response 87121604

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050523\
Data File : PD075307.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2023 02:58
Operator : AR\AJ
Sample : 02479-15
Misc :
ALS Vial : 45 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
EW974

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/09/2023
Supervised By :Ankita Jodhani 05/09/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 06 06:10:38 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(17) 4,4'-DDT (MA)

7.012min 80.307 ng/ml m

response 203839870

(17) 4,4'-DDT #2 (MA)

6.031min 86.441 ng/ml

response 87121604