

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050621\
 Data File : PD062735.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 May 2021 19:38
 Operator : AR\AJ
 Sample : PSTDICV050
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 ICVPD050621

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 06 08:52:11 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050621.M
 Quant Title : GC Extractables
 QLast Update : Thu May 06 08:49:38 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.442	3.971	44893678	584.6E6	50.487	50.497
2)	SA Decachlor...	8.202	9.031	53927449	586.7E6	50.937	50.842
Target Compounds							
2)	A alpha-BHC	3.878	4.360	71502007	963.2E6	51.418	50.032
3)	MA gamma-BHC...	4.162	4.645	66297781	871.9E6	50.736	50.447
4)	MA Heptachlor	4.463	5.153	64774909	876.3E6	51.022	49.926
5)	MB Aldrin	4.712	5.457	61661866	840.5E6	51.214	50.254
6)	B beta-BHC	4.413	4.815	30036682	380.0E6	50.135	49.826
7)	B delta-BHC	4.617	5.032	69618338	898.4E6	51.632	49.769
8)	B Heptachlo...	5.155	5.844	58690376	779.4E6	50.723	48.788
9)	A Endosulfan I	5.494	6.200	55669514	738.4E6	50.818	50.026
10)	B gamma-Chl...	5.382	6.078	59792472	801.4E6	50.800	50.370
11)	B alpha-Chl...	5.440	6.150	58699526	766.5E6	50.683	50.212
12)	B 4,4'-DDE	5.604	6.308	58496055	791.0E6	50.870	49.942
13)	MA Dieldrin	5.738	6.457	61417549	803.9E6	51.143	49.757
14)	MA Endrin	5.995	6.675	54798167	682.2E6	49.306	48.524
15)	B Endosulfa...	6.269	6.884	53007442	668.5E6	50.882	50.038
16)	A 4,4'-DDD	6.121	6.797	49446572	653.1E6	51.927	50.540
17)	MA 4,4'-DDT	6.362	7.096	50622408	648.6E6	52.728	50.915
18)	B Endrin al...	6.437	7.008	46977733	584.9E6	50.358	50.130
19)	B Endosulfa...	6.651	7.231	54224224	700.5E6	51.787	50.368
20)	A Methoxychlor	6.910	7.555	30297391	333.6E6	51.467	49.977
21)	B Endrin ke...	7.143	7.702	65229651	692.7E6	52.323	51.046
22)	Mirex	7.332	8.156	47743950	457.2E6	52.152	49.912

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050621\
Data File : PD062735.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 May 2021 19:38
Operator : AR\AJ
Sample : PSTDICV050
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
ICVPD050621

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 06 08:52:11 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050621.M
Quant Title : GC Extractables
QLast Update : Thu May 06 08:49:38 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

