

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050720\
 Data File : PD057868.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 May 2020 12:06
 Operator : AJ\MA
 Sample : INDA306
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA306

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 08 05:09:00 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD042920CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Apr 30 07:30:11 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.229	3.896	15420444	26016003	18.468	18.947
27)	SA Decachlor...	7.876	8.902	33146308	48908482	34.483	33.777
Target Compounds							
2)	A alpha-BHC	3.652	4.280	22397762	37250977	19.254	19.607
3)	MA gamma-BHC...	3.924	4.560	21490327	35460979	18.332	19.630
4)	MA Heptachlor	4.204	5.062	17761115	35284046	17.106	18.960
9)	A Endosulfan I	5.202	6.100	17309450	30174781	19.299	19.282
13)	MA Dieldrin	5.441	6.355	38569354	65634572	37.568	38.555
14)	MA Endrin	5.692	6.571	31221528	51771052	34.713	34.128
16)	A 4,4'-DDD	5.830	6.697	29327450	51445262	36.089	38.431
17)	MA 4,4'-DDT	6.063	6.995	27961963	51210859	34.064	36.051
20)	A Methoxychlor	6.612	7.454	64645255	131.2E6	150.225	169.572

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050720\
Data File : PD057868.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 May 2020 12:06
Operator : AJ\MA
Sample : INDA306
Misc :
ALS Vial : 8 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 08 05:09:00 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD042920CLP.M
Quant Title : GC Extractables
QLast Update : Thu Apr 30 07:30:11 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

