

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD050722CLP.M
 Title : GC Extractables
 Last Update : Mon May 09 06:13:38 2022
 Response Via : Initial Calibration

Calibration Files

3 =PD069462.D 1 =PD069460.D 2 =PD069461.D
 4 =PD069463.D 5 =PD069464.D 6 =PD069465.D

Compound	3	1	2	4	5	6	Avg	%RSD
1) SA Tetrachloro-m...	1.106	1.271	1.186	1.084	1.096	1.082	1.137 E6	6.66
2) A alpha-BHC							0.000	-1.00
3) MA gamma-BHC (Li...							0.000	-1.00
4) MA Heptachlor							0.000	-1.00
5) MB Aldrin							0.000	-1.00
6) B beta-BHC							0.000	-1.00
7) B delta-BHC							0.000	-1.00
8) B Heptachlor ep...							0.000	-1.00
9) A Endosulfan I							0.000	-1.00
10) B trans-Chlordane							0.000	-1.00
11) B cis-Chlordane							0.000	-1.00
12) B 4,4'-DDE							0.000	-1.00
13) MA Dieldrin							0.000	-1.00
14) MA Endrin							0.000	-1.00
15) B Endosulfan II							0.000	-1.00
16) A 4,4'-DDD							0.000	-1.00
17) MA 4,4'-DDT							0.000	-1.00
18) B Endrin aldehyde							0.000	-1.00
19) B Endosulfan Su...							0.000	-1.00
20) A Methoxychlor							0.000	-1.00
21) B Endrin ketone							0.000	-1.00
22) Toxaphene-1	7.519	8.814	8.078	7.382	7.543	7.408	7.791 E3	7.21
23) Toxaphene-2	7.748	8.462	8.243	8.096	8.309	8.670	8.255 E3	3.83
24) Toxaphene-3	8.182	9.472	8.759	8.484	8.762	8.662	8.720 E3	4.91
25) Toxaphene-4	3.498	3.606	3.541	3.653	3.741	3.708	3.625 E4	2.60
26) Toxaphene-5	1.633	1.750	1.637	1.724	1.790	1.829	1.727 E4	4.62
27) SA Decachlorobip...	1.059	1.335	1.165	1.006	0.996	0.996	1.093 E6	12.37

Signal #2 Calibration Files

3 =PD069462.D 1 =PD069460.D 2 =PD069461.D
 4 =PD069463.D 5 =PD069464.D 6 =PD069465.D

Compound	3	1	2	4	5	6	Avg	%RSD
1) SA Tetrachloro-m...	1.154	1.358	1.253	1.086	1.043	0.987	1.147 E7	12.08
2) A alpha-BHC							0.000	-1.00
3) MA gamma-BHC (Li...							0.000	-1.00
4) MA Heptachlor							0.000	-1.00
5) MB Aldrin							0.000	-1.00
6) B beta-BHC							0.000	-1.00
7) B delta-BHC							0.000	-1.00
8) B Heptachlor ep...							0.000	-1.00
9) A Endosulfan I							0.000	-1.00
10) B trans-Chlordane							0.000	-1.00
11) B cis-Chlordane							0.000	-1.00
12) B 4,4'-DDE							0.000	-1.00
13) MA Dieldrin							0.000	-1.00
14) MA Endrin							0.000	-1.00
15) B Endosulfan II							0.000	-1.00
16) A 4,4'-DDD							0.000	-1.00
17) MA 4,4'-DDT							0.000	-1.00
18) B Endrin aldehyde							0.000	-1.00
19) B Endosulfan Su...							0.000	-1.00
20) A Methoxychlor							0.000	-1.00
21) B Endrin ketone							0.000	-1.00
22) Toxaphene-1	0.874	1.063	0.969	0.820	0.768	0.723	0.869 E5	14.69
23) Toxaphene-2	0.985	1.207	1.076	0.906	0.864	0.820	0.976 E5	14.86

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD050722CLP.M
 Title : GC Extractables
 Last Update : Mon May 09 06:13:38 2022
 Response Via : Initial Calibration

Calibration Files

3 =PD069462.D 1 =PD069460.D 2 =PD069461.D
 4 =PD069463.D 5 =PD069464.D 6 =PD069465.D

	Compound	3	1	2	4	5	6	Avg	%RSD
24)	Toxaphene-3	2.523	2.961	2.722	2.408	2.329	2.250	2.532 E5	10.54
25)	Toxaphene-4	1.948	2.309	2.119	1.854	1.817	1.757	1.967 E5	10.66
26)	Toxaphene-5	1.255	1.455	1.298	1.167	1.168	1.166	1.252 E5	9.09
27) SA	Decachlorobip...	6.672	8.446	7.472	6.094	5.735	5.371	6.632 E6	17.43

(#) = Out of Range