

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD050722CLP.M
 Title : GC Extractables
 Last Update : Sat May 07 05:37:19 2022
 Response Via : Initial Calibration

Calibration Files

3 =PD069470.D 1 =PD069466.D 2 =PD069468.D
 4 =PD069472.D 5 =PD069474.D

	Compound	3	1	2	4	5	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.221	1.320	1.231	1.165	1.141	1.216 E6	5.72
2) A	alpha-BHC	1.706	1.712	1.640	1.717	1.753	1.706 E6	2.40
3) MA	gamma-BHC (Lindane)	1.663	1.796	1.663	1.620	1.630	1.674 E6	4.22
4) MA	Heptachlor	1.614	1.722	1.606	1.578	1.569	1.618 E6	3.78
5) MB	Aldrin						0.000	-1.00
6) B	beta-BHC						0.000	-1.00
7) B	delta-BHC						0.000	-1.00
8) B	Heptachlor epoxide						0.000	-1.00
9) A	Endosulfan I	1.377	1.480	1.381	1.322	1.317	1.375 E6	4.76
10) B	trans-Chlordane						0.000	-1.00
11) B	cis-Chlordane						0.000	-1.00
12) B	4,4'-DDE						0.000	-1.00
13) MA	Dieldrin	1.484	1.523	1.450	1.469	1.488	1.483 E6	1.83
14) MA	Endrin	1.314	1.336	1.271	1.299	1.290	1.302 E6	1.88
15) B	Endosulfan II						0.000	-1.00
16) A	4,4'-DDD	1.176	1.208	1.145	1.163	1.168	1.172 E6	1.99
17) MA	4,4'-DDT	1.187	1.141	1.119	1.189	1.210	1.169 E6	3.22
18) B	Endrin aldehyde						0.000	-1.00
19) B	Endosulfan Sulfate						0.000	-1.00
20) A	Methoxychlor	6.662	7.136	6.664	6.361	6.022	6.569 E5	6.28
21) B	Endrin ketone						0.000	-1.00
22)	Toxaphene-1						0.000	-1.00
23)	Toxaphene-2						0.000	-1.00
24)	Toxaphene-3						0.000	-1.00
25)	Toxaphene-4						0.000	-1.00
26)	Toxaphene-5						0.000	-1.00
27) SA	Decachlorobiphenyl	1.249	1.431	1.281	1.146	1.094	1.240 E6	10.55

Signal #2 Calibration Files

3 =PD069470.D 1 =PD069466.D 2 =PD069468.D
 4 =PD069472.D 5 =PD069474.D

	Compound	3	1	2	4	5	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.277	1.448	1.327	1.181	1.107	1.268 E7	10.40
2) A	alpha-BHC	1.831	2.056	1.880	1.720	1.645	1.826 E7	8.64
3) MA	gamma-BHC (Lindane)	1.641	1.824	1.682	1.528	1.466	1.628 E7	8.57
4) MA	Heptachlor	1.604	1.787	1.648	1.475	1.397	1.582 E7	9.62
5) MB	Aldrin						0.000	-1.00
6) B	beta-BHC						0.000	-1.00
7) B	delta-BHC						0.000	-1.00
8) B	Heptachlor epoxide						0.000	-1.00
9) A	Endosulfan I	1.276	1.449	1.336	1.158	1.099	1.264 E7	11.07
10) B	trans-Chlordane						0.000	-1.00
11) B	cis-Chlordane						0.000	-1.00
12) B	4,4'-DDE						0.000	-1.00
13) MA	Dieldrin	1.286	1.422	1.316	1.200	1.142	1.273 E7	8.49
14) MA	Endrin	1.027	1.119	1.031	0.972	0.909	1.011 E7	7.70
15) B	Endosulfan II						0.000	-1.00
16) A	4,4'-DDD	1.000	1.115	1.029	0.928	0.877	0.990 E7	9.29
17) MA	4,4'-DDT	0.973	1.039	0.984	0.918	0.869	0.957 E7	6.79
18) B	Endrin aldehyde						0.000	-1.00
19) B	Endosulfan Sulfate						0.000	-1.00
20) A	Methoxychlor	4.449	5.115	4.656	4.101	3.720	4.408 E6	12.06
21) B	Endrin ketone						0.000	-1.00
22)	Toxaphene-1						0.000	-1.00
23)	Toxaphene-2						0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
Method File : PD050722CLP.M
Title : GC Extractables
Last Update : Sat May 07 05:37:19 2022
Response Via : Initial Calibration

Calibration Files

3 =PD069470.D 1 =PD069466.D 2 =PD069468.D
4 =PD069472.D 5 =PD069474.D

	Compound	3	1	2	4	5	Avg	%RSD
24)	Toxaphene-3						0.000	-1.00
25)	Toxaphene-4						0.000	-1.00
26)	Toxaphene-5						0.000	-1.00
27) SA	Decachlorobiphenyl	7.403	8.591	7.708	6.693	6.118	7.303 E6	13.01

(#) = Out of Range