

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050821\
 Data File : PD062824.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 May 2021 19:42
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 08 08:04:22 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050621.M
 Quant Title : GC Extractables
 QLast Update : Thu May 06 08:49:38 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.441	3.971	48244180	585.4E6	54.255	50.569
2)	SA Decachlor...	8.202	9.030	58119543	612.2E6	54.897	53.052
Target Compounds							
2)	A alpha-BHC	3.878	4.360	77518349	969.8E6	55.744	50.371
3)	MA gamma-BHC...	4.162	4.644	71700159	876.2E6	54.870	50.694
4)	MA Heptachlor	4.463	5.152	68913121	876.8E6	54.282	49.952
5)	MB Aldrin	4.711	5.457	67880499	840.1E6	56.380	50.233
6)	B beta-BHC	4.412	4.815	31557584	382.1E6	52.673	50.102
7)	B delta-BHC	4.617	5.031	75108350	922.2E6	55.704	51.090
8)	B Heptachlo...	5.154	5.843	63433813	831.2E6	54.822	52.036
9)	A Endosulfan I	5.494	6.200	59782212	736.6E6	54.572	49.902
10)	B gamma-Chl...	5.382	6.077	64666370	800.5E6	54.940	50.309
11)	B alpha-Chl...	5.441	6.149	63221418	759.6E6	54.588	49.757
12)	B 4,4'-DDE	5.604	6.307	65782992	783.2E6	57.207	49.449
13)	MA Dieldrin	5.737	6.456	66273842	820.4E6	55.187	50.772
14)	MA Endrin	5.995	6.674	57700444	706.6E6	51.918	50.255
15)	B Endosulfa...	6.269	6.883	56746454	670.5E6	54.471	50.188
16)	A 4,4'-DDD	6.121	6.796	54006680	658.2E6	56.716	50.937
17)	MA 4,4'-DDT	6.362	7.095	52761648	635.8E6	54.957	49.911
18)	B Endrin al...	6.437	7.007	49561665	580.3E6	53.127	49.737
19)	B Endosulfa...	6.651	7.231	58027265	709.2E6	55.419	50.994
20)	A Methoxychlor	6.909	7.553	31153663	322.6E6	52.921	48.328
21)	B Endrin ke...	7.143	7.701	71228168	709.4E6	57.135	52.273
22)	Mirex	7.331	8.155	51774585	473.8E6	56.555	51.723

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050821\
Data File : PD062824.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 May 2021 19:42
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 08 08:04:22 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050621.M
Quant Title : GC Extractables
QLast Update : Thu May 06 08:49:38 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

