

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050923\
 Data File : PD075372.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 May 2023 14:36
 Operator : AR\AJ
 Sample : 02419-25DL 6X
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 EW922DL

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 05/10/2023
 Supervised By : Ankita Jodhani 05/14/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 09 21:15:31 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu May 04 02:14:07 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.549	2.771	3613862	1727156	1.584	1.638
27)	SA Decachlor...	9.083	7.918	6606040	3996901	2.669m	4.232 #
Target Compounds							
4)	MA Heptachlor	4.929	3.942	8644820	4929565	2.319m	3.374m#
5)	MB Aldrin	5.271	4.223	16060153	6845477	4.403m	4.616m
8)	B Heptachlo...	5.699	4.726	20831945	5759737	6.146	4.130 #
10)	B trans-Chl...	5.957	4.978	121.9E6	50730504	37.206m	37.177
11)	B cis-Chlor...	6.040	5.041	89570417	29072782	26.282	20.832
12)	B 4,4'-DDE	6.208	5.231	15421142	8930651	4.722	6.947 #
13)	MA Dieldrin	6.363	5.361	434.0E6	181.6E6	126.898m	133.074m
14)	MA Endrin	6.594	5.639	15504684	9686485	5.362	7.964 #
16)	A 4,4'-DDD	6.727	5.785	19478435	8067329	8.060	8.137
17)	MA 4,4'-DDT	7.041	6.036	16812590	8697511	6.624m	8.630 #
21)	B Endrin ke...	7.664	6.840	3369655	3140599	1.112m	2.382 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050923\
Data File : PD075372.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 May 2023 14:36
Operator : AR\AJ
Sample : 02419-25DL 6X
Misc :
ALS Vial : 13 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 09 21:15:31 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Manual Integrations
APPROVED
Reviewed By: Abdul Mirza 05/10/2023
Supervised By: Ankita Jodhani 05/14/2023

Instrument :
ECD_D
Client Sampled :
EW922DL