

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050923\
 Data File : PD075373.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 May 2023 15:37
 Operator : AR\AJ
 Sample : 02419-24DL 2X
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 EW921DL

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 05/10/2023
 Supervised By : Ankita Jodhani 05/14/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 09 21:16:21 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu May 04 02:14:07 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.553	2.770	25611840	8261713	11.229	7.833 #
2)	SA Decachlor...	9.100	7.917	66550499	49226101	26.892m	52.125 #
Target Compounds							
3)	MA gamma-BHC...	4.350	3.581	196.0E6	55686600	54.525m	37.884m#
14)	MA Endrin	6.599	5.608	2162.4E6	873.1E6	747.789m	717.794m
17)	MA 4,4'-DDT	7.023	6.033	2265.5E6	916.2E6	892.547m	909.064m
19)	B Endosulfa...	7.167	6.333	127.4E6	65691181	46.314	56.483m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050923\
Data File : PD075373.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 May 2023 15:37
Operator : AR\AJ
Sample : 02419-24DL 2X
Misc :
ALS Vial : 14 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 09 21:16:21 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Reviewed By: Abdul Mirza 05/10/2023
Supervised By: Ankita Jodhani 05/14/2023

Manual Integrations
APPROVED

Instrument :
ECD_D
Client Sampled :
EW921DL