

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050925\
 Data File : PD088448.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 May 2025 16:19
 Operator : AR\AJ
 Sample : TOXAPH501
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 TOXAPH5024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 09 17:11:17 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050925CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri May 09 16:32:44 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.552	2.883	188.9E6	1216.7E6	80.926	74.121
27)	SA Decachlor...	9.077	8.077	458.4E6	2331.7E6	149.605	148.147
Target Compounds							
22)	Toxaphene-1	6.244	5.150	215.8E6	802.4E6	8467.503	7292.927
23)	Toxaphene-2	6.444	6.122	315.4E6	898.3E6	8739.475	7338.838
24)	Toxaphene-3	7.059	6.760	889.2E6	3044.1E6	9124.401	8050.109
25)	Toxaphene-4	7.151	7.202	667.5E6	1966.5E6	9361.757	7487.860
26)	Toxaphene-5	7.932	7.333	466.4E6	1583.2E6	9001.889	8328.376

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050925\
Data File : PD088448.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 May 2025 16:19
Operator : AR\AJ
Sample : TOXAPH501
Misc :
ALS Vial : 8 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 09 17:11:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050925CLP.M
Quant Title : GC Extractables
QLast Update : Fri May 09 16:32:44 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

