

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050925\
 Data File : PD088455.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 May 2025 17:55
 Operator : AR\AJ
 Sample : INDA401
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA4031

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 12 11:50:19 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050925CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon May 12 11:50:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.551	2.882	87160796	592.8E6	39.437	39.677
2)	SA Decachlor...	9.076	8.076	279.7E6	1414.3E6	78.568	79.082
Target Compounds							
2)	A alpha-BHC	4.001	3.395	196.6E6	961.2E6	39.389	39.702
3)	MA gamma-BHC...	4.332	3.732	187.6E6	883.4E6	39.465	39.755
4)	MA Heptachlor	4.931	4.085	188.5E6	895.8E6	39.343	39.590
9)	A Endosulfan I	6.076	5.250	153.5E6	747.1E6	39.375	39.757
13)	MA Dieldrin	6.348	5.515	348.0E6	1604.7E6	79.413	79.626
14)	MA Endrin	6.576	5.792	288.0E6	1435.0E6	78.880	79.677
16)	A 4,4'-DDD	6.707	5.932	240.6E6	1312.8E6	78.900	79.527
17)	MA 4,4'-DDT	7.023	6.187	253.5E6	1368.4E6	78.795	79.675
20)	A Methoxychlor	7.495	6.757	662.6E6	2889.5E6	395.461	399.074

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050925\
Data File : PD088455.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 May 2025 17:55
Operator : AR\AJ
Sample : INDA401
Misc :
ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 12 11:50:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050925CLP.M
Quant Title : GC Extractables
QLast Update : Mon May 12 11:50:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

