

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051021\
 Data File : PD062842.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 May 2021 11:36
 Operator : AR\AJ
 Sample : M2296-01MSD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 SP-1MSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 10 13:45:47 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050621.M
 Quant Title : GC Extractables
 QLast Update : Thu May 06 08:49:38 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.442	3.971	19357123	233.4E6	21.769	20.157
28)	SA Decachlor...	8.201	9.030	21770128	238.0E6	20.563	20.630
Target Compounds							
2)	A alpha-BHC	3.879	4.359	74549573	909.7E6	53.609	47.250
3)	MA gamma-BHC...	4.162	4.644	69021532	842.0E6	52.820	48.715
4)	MA Heptachlor	4.463	5.152	67353599	859.8E6	53.053	48.985
5)	MB Aldrin	4.712	5.456	65522868	816.9E6	54.421	48.846
6)	B beta-BHC	4.413	4.814	29321523	357.2E6	48.941	46.842
7)	B delta-BHC	4.617	5.031	70407751	775.9E6	52.218	42.984
8)	B Heptachlo...	5.154	5.842	60822733	733.9E6	52.566	45.945
9)	A Endosulfan I	5.494	6.198	57032446	693.3E6	52.062	46.970
10)	B gamma-Chl...	5.381	6.076	63525821	723.6E6	53.971	45.479
11)	B alpha-Chl...	5.440	6.149	61171546	752.0E6	52.818	49.261
12)	B 4,4'-DDE	5.604	6.306	73632837	891.6E6	64.034	56.293
13)	MA Dieldrin	5.737	6.455	61456448	755.0E6	51.175	46.726
14)	MA Endrin	5.995	6.674	57403253	639.4E6	51.650	45.480
15)	B Endosulfa...	6.268	6.883	51986981	635.7E6	49.902	47.588
16)	A 4,4'-DDD	6.121	6.795	50955156	695.0E6	53.511	53.789
17)	MA 4,4'-DDT	6.362	7.095	57272320	702.0E6	59.655	55.111
18)	B Endrin al...	6.437	7.007	43232729	493.9E6	46.343	42.333
19)	B Endosulfa...	6.651	7.230	52160648	631.3E6	49.816	45.390
20)	A Methoxychlor	6.909	7.553	28110678	304.1E6	47.752	45.564
21)	B Endrin ke...	7.143	7.700	64020158	646.7E6	51.353	47.656
22)	Mirex	7.330	8.154	46308248	425.6E6	50.584	46.459

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051021\
Data File : PD062842.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 May 2021 11:36
Operator : AR\AJ
Sample : M2296-01MSD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
SP-1MSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 10 13:45:47 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050621.M
Quant Title : GC Extractables
QLast Update : Thu May 06 08:49:38 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

