

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051023\
 Data File : PD075409.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 May 2023 22:28
 Operator : AR\AJ
 Sample : INDA416
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA3243

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 05/12/2023
 Supervised By : Ankita Jodhani 05/12/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 11 11:29:38 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu May 04 02:14:07 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.548	2.770	46282966	21978413	20.292	20.838
27) SA Decachlor...	9.088	7.916	85453055	33839700	34.530	35.833
Target Compounds						
2) A alpha-BHC	4.004	3.273	73344331	30917451	19.072	19.960
3) MA gamma-BHC...	4.338	3.603	68450749	28794082	19.040	19.589
4) MA Heptachlor	4.929	3.943	68785605	28099532	18.451	19.231
9) A Endosulfan I	6.084	5.095	56169170	23290595	18.621m	19.406m
13) MA Dieldrin	6.361	5.361	126.4E6	51240483	36.972	37.556
14) MA Endrin	6.592	5.637	100.3E6	41738840	34.698	34.316
16) A 4,4'-DDD	6.725	5.785	100.5E6	41813042	41.569	42.177
17) MA 4,4'-DDT	7.040	6.036	76291068	31321883	30.057	31.077
20) A Methoxychlor	7.516	6.612	231.6E6	92400760	157.653	161.319

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051023\
Data File : PD075409.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 May 2023 22:28
Operator : AR\AJ
Sample : INDA416
Misc :
ALS Vial : 13 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 11 11:29:38 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

