

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051023\
 Data File : PD075434.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 May 2023 04:06
 Operator : AR\AJ
 Sample : 02694-21MS
 Misc :
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 JREQ1MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 11 05:17:16 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu May 04 02:14:07 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.548	2.771	41937601	19479579	18.387	18.469
27)	SA Decachlor...	9.088	7.915	71923022	28228487	29.063	29.891
Target Compounds							
2)	A alpha-BHC	4.005	3.273	173.6E6	68452223	45.144	44.191
3)	MA gamma-BHC...	4.337	3.603	149.9E6	60941664	41.684	41.459
4)	MA Heptachlor	4.929	3.943	148.1E6	57324060	39.719	39.231
5)	MB Aldrin	5.272	4.223	155.1E6	59684166	42.509	40.243
6)	B beta-BHC	4.533	3.899	63669954	27047786	38.788	37.146
7)	B delta-BHC	4.782	4.129	121.7E6	49562333	36.013	31.367
8)	B Heptachlo...	5.699	4.726	139.9E6	54071377	41.272	38.767
9)	A Endosulfan I	6.086	5.096	128.9E6	50064008	42.720	41.714
10)	B trans-Chl...	5.956	4.976	136.9E6	53940660	41.792	39.530
11)	B cis-Chlor...	6.035	5.040	139.6E6	54880991	40.959	39.325
12)	B 4,4'-DDE	6.207	5.230	273.6E6	103.5E6	83.784	80.548
13)	MA Dieldrin	6.361	5.361	302.7E6	114.9E6	88.509	84.209
14)	MA Endrin	6.592	5.636	238.3E6	96806999	82.417	79.590
15)	B Endosulfa...	6.811	5.930	249.3E6	94128010	77.419	73.567
16)	A 4,4'-DDD	6.725	5.784	234.1E6	90758999	96.889	91.548
17)	MA 4,4'-DDT	7.040	6.035	184.8E6	70615739	72.813	70.064
18)	B Endrin al...	6.941	6.110	176.1E6	68484676	74.453	67.441
19)	B Endosulfa...	7.176	6.333	214.0E6	84775660	77.810	72.892
20)	A Methoxychlor	7.515	6.612	532.0E6	197.8E6	362.150	345.258
21)	B Endrin ke...	7.662	6.840	261.0E6	105.5E6	86.120	79.984

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051023\
Data File : PD075434.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 May 2023 04:06
Operator : AR\AJ
Sample : 02694-21MS
Misc :
ALS Vial : 50 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 11 05:17:16 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

