

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051123\
 Data File : PD075448.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 May 2023 15:41
 Operator : AR\AJ
 Sample : INDA417
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA3235

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 11 21:25:17 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu May 04 02:14:07 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.547	2.769	42290867	20794237	18.542	19.715
27)	SA Decachlor...	9.088	7.916	89301292	33864828	36.085	35.859
Target Compounds							
2)	A alpha-BHC	4.003	3.272	69888491	30123425	18.174	19.447
3)	MA gamma-BHC...	4.337	3.602	64939143	28078103	18.063	19.102
4)	MA Heptachlor	4.928	3.942	65207381	28127479	17.492	19.250
9)	A Endosulfan I	6.085	5.096	52712281	22756464	17.475	18.961
13)	MA Dieldrin	6.361	5.361	122.4E6	51038459	35.803	37.408
14)	MA Endrin	6.591	5.637	102.6E6	43829264	35.475	36.034
16)	A 4,4'-DDD	6.724	5.784	91809190	38441014	37.992	38.775
17)	MA 4,4'-DDT	7.040	6.036	84694579	34123699	33.367	33.857
20)	A Methoxychlor	7.515	6.611	263.8E6	104.8E6	179.577	182.904

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051123\
Data File : PD075448.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 May 2023 15:41
Operator : AR\AJ
Sample : INDA417
Misc :
ALS Vial : 9 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 11 21:25:17 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

