

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051220\
 Data File : PD057897.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 May 2020 16:05
 Operator : AJ\MA
 Sample : L2575-02
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 TP-28

Manual Integrations
 APPROVED

mohammad
 5/13/2020 11:19:14 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 13 06:47:37 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051220.M
 Quant Title : GC Extractables
 QLast Update : Tue May 12 12:00:03 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.228	3.896	18991809	25626460	21.921	21.785
28)	SA Decachlor...	7.873	8.899	19814863	24328836	21.992	21.350m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051220\
Data File : PD057897.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 May 2020 16:05
Operator : AJ\MA
Sample : L2575-02
Misc :
ALS Vial : 15 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleID :
TP-28


Manual Integrations
APPROVED

mohammad
5/13/2020 11:19:14 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 13 06:47:37 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051220.M
Quant Title : GC Extractables
QLast Update : Tue May 12 12:00:03 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene

R.T.: 3.228 min
Delta R.T.: 0.000 min
Response: 18991809
Conc: 21.92 ng/ml

Instrument :
ECD_D
ClientSampleId :
TP-28


Manual Integrations
APPROVED

mohammad
5/13/2020 11:19:14 AM




#1 Tetrachloro-m-xylene

R.T.: 3.896 min
Delta R.T.: 0.000 min
Response: 25626460
Conc: 21.78 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.873 min
Delta R.T.: 0.000 min
Response: 19814863
Conc: 21.99 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.899 min
Delta R.T.: -0.002 min
Response: 24328836
Conc: 21.35 ng/ml m