

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051225\
 Data File : PD088484.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 May 2025 19:35
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 13 01:29:37 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 19 06:32:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.551	2.883	98799964	652.8E6	49.464	44.646
28)	SA Decachlor...	9.073	8.074	160.6E6	768.5E6	48.547	41.589
Target Compounds							
2)	A alpha-BHC	4.000	3.395	222.0E6	1035.0E6	51.485	45.274
3)	MA gamma-BHC...	4.331	3.732	213.5E6	964.7E6	50.996	44.789
4)	MA Heptachlor	4.931	4.085	212.5E6	963.3E6	52.613	45.019
5)	MB Aldrin	5.272	4.371	206.2E6	952.4E6	52.205	45.782
6)	B beta-BHC	4.516	4.027	82615547	419.9E6	50.891	45.377
7)	B delta-BHC	4.764	4.264	212.4E6	964.2E6	51.483	45.347
8)	B Heptachlo...	5.692	4.875	185.2E6	852.4E6	51.821	45.114
9)	A Endosulfan I	6.075	5.249	176.6E6	790.7E6	52.307	43.898
10)	B gamma-Chl...	5.946	5.127	192.1E6	913.5E6	53.040	45.008
11)	B alpha-Chl...	6.028	5.192	189.8E6	879.0E6	52.576	44.838
12)	B 4,4'-DDE	6.196	5.377	168.0E6	901.5E6	50.921	45.775
13)	MA Dieldrin	6.348	5.515	187.6E6	899.2E6	52.567	45.110
14)	MA Endrin	6.575	5.791	151.1E6	801.7E6	50.664	44.044
15)	B Endosulfa...	6.786	6.082	157.9E6	778.0E6	50.507	44.401
16)	A 4,4'-DDD	6.705	5.931	130.9E6	739.7E6	52.059	45.141
17)	MA 4,4'-DDT	7.021	6.185	143.4E6	771.9E6	51.625	45.364
18)	B Endrin al...	6.915	6.260	116.3E6	577.7E6	50.389	43.434
19)	B Endosulfa...	7.149	6.483	148.1E6	754.0E6	51.513	44.008
20)	A Methoxychlor	7.493	6.756	75695695	400.9E6	50.479	43.950
21)	B Endrin ke...	7.630	6.993	158.8E6	830.6E6	51.460	44.406
22)	Mirex	8.114	7.187	118.9E6	642.0E6	50.644	43.319



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051225\
Data File : PD088484.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 May 2025 19:35
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 13 01:29:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041825.M
Quant Title : GC Extractables
QLast Update : Sat Apr 19 06:32:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.551 min
Delta R.T.: 0.000 min
Response: 98799964
Conc: 49.46 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#1 Tetrachloro-m-xylene

R.T.: 2.883 min
Delta R.T.: 0.000 min
Response: 652794051
Conc: 44.65 ng/ml




#2 alpha-BHC

R.T.: 4.000 min
Delta R.T.: 0.000 min
Response: 222002621
Conc: 51.48 ng/ml



#2 alpha-BHC


R.T.: 3.395 min
Delta R.T.: 0.000 min
Response: 1035030981
Conc: 45.27 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.331 min
Delta R.T.: 0.000 min
Response: 213545788
Conc: 51.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#3 gamma-BHC (Lindane)

R.T.: 3.732 min
Delta R.T.: -0.001 min
Response: 964687763
Conc: 44.79 ng/ml




#4 Heptachlor

R.T.: 4.931 min
Delta R.T.: 0.000 min
Response: 212502219
Conc: 52.61 ng/ml



#4 Heptachlor


R.T.: 4.085 min
Delta R.T.: -0.001 min
Response: 963326787
Conc: 45.02 ng/ml



#5 Aldrin


R.T.: 5.272 min
Delta R.T.: -0.001 min
Response: 206187455
Conc: 52.20 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#5 Aldrin

R.T.: 4.371 min
Delta R.T.: -0.002 min
Response: 952392906
Conc: 45.78 ng/ml




#6 beta-BHC

R.T.: 4.516 min
Delta R.T.: 0.000 min
Response: 82615547
Conc: 50.89 ng/ml



#6 beta-BHC


R.T.: 4.027 min
Delta R.T.: -0.001 min
Response: 419926658
Conc: 45.38 ng/ml



#7 delta-BHC


R.T.: 4.764 min
Delta R.T.: -0.001 min
Response: 212364691
Conc: 51.48 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#7 delta-BHC

R.T.: 4.264 min
Delta R.T.: -0.001 min
Response: 964158309
Conc: 45.35 ng/ml




#8 Heptachlor epoxide

R.T.: 5.692 min
Delta R.T.: 0.000 min
Response: 185206441
Conc: 51.82 ng/ml



#8 Heptachlor epoxide


R.T.: 4.875 min
Delta R.T.: -0.002 min
Response: 852395183
Conc: 45.11 ng/ml



#9 Endosulfan I


R.T.: 6.075 min
Delta R.T.: 0.000 min
Response: 176648769
Conc: 52.31 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#9 Endosulfan I

R.T.: 5.249 min
Delta R.T.: -0.002 min
Response: 790704067
Conc: 43.90 ng/ml




#10 gamma-Chlordane

R.T.: 5.946 min
Delta R.T.: 0.000 min
Response: 192115063
Conc: 53.04 ng/ml



#10 gamma-Chlordane


R.T.: 5.127 min
Delta R.T.: -0.002 min
Response: 913483300
Conc: 45.01 ng/ml



#11 alpha-Chlordane


R.T.: 6.028 min
Delta R.T.: 0.000 min
Response: 189776187
Conc: 52.58 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#11 alpha-Chlordane

R.T.: 5.192 min
Delta R.T.: -0.002 min
Response: 878951558
Conc: 44.84 ng/ml




#12 4,4'-DDE

R.T.: 6.196 min
Delta R.T.: -0.001 min
Response: 167950618
Conc: 50.92 ng/ml



#12 4,4'-DDE


R.T.: 5.377 min
Delta R.T.: -0.003 min
Response: 901484116
Conc: 45.77 ng/ml



#13 Dieldrin


R.T.: 6.348 min
Delta R.T.: 0.000 min
Response: 187586813
Conc: 52.57 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#13 Dieldrin

R.T.: 5.515 min
Delta R.T.: -0.002 min
Response: 899168284
Conc: 45.11 ng/ml




#14 Endrin

R.T.: 6.575 min
Delta R.T.: 0.000 min
Response: 151107861
Conc: 50.66 ng/ml



#14 Endrin


R.T.: 5.791 min
Delta R.T.: -0.002 min
Response: 801695321
Conc: 44.04 ng/ml



#15 Endosulfan II


R.T.: 6.786 min
Delta R.T.: 0.000 min
Response: 157894304
Conc: 50.51 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#15 Endosulfan II

R.T.: 6.082 min
Delta R.T.: -0.003 min
Response: 777984241
Conc: 44.40 ng/ml




#16 4,4'-DDD

R.T.: 6.705 min
Delta R.T.: 0.000 min
Response: 130927094
Conc: 52.06 ng/ml



#16 4,4'-DDD


R.T.: 5.931 min
Delta R.T.: -0.003 min
Response: 739693615
Conc: 45.14 ng/ml



#17 4,4'-DDT


R.T.: 7.021 min
Delta R.T.: 0.000 min
Response: 143391903
Conc: 51.63 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#17 4,4'-DDT

R.T.: 6.185 min
Delta R.T.: -0.003 min
Response: 771939735
Conc: 45.36 ng/ml




#18 Endrin aldehyde

R.T.: 6.915 min
Delta R.T.: 0.000 min
Response: 116296880
Conc: 50.39 ng/ml



#18 Endrin aldehyde


R.T.: 6.260 min
Delta R.T.: -0.003 min
Response: 577709895
Conc: 43.43 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.149 min
Delta R.T.: 0.000 min
Response: 148144110
Conc: 51.51 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#19 Endosulfan Sulfate

R.T.: 6.483 min
Delta R.T.: -0.003 min
Response: 753997833
Conc: 44.01 ng/ml




#20 Methoxychlor

R.T.: 7.493 min
Delta R.T.: -0.002 min
Response: 75695695
Conc: 50.48 ng/ml



#20 Methoxychlor


R.T.: 6.756 min
Delta R.T.: -0.003 min
Response: 400867812
Conc: 43.95 ng/ml



#21 Endrin ketone


R.T.: 7.630 min
Delta R.T.: 0.000 min
Response: 158836382
Conc: 51.46 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#21 Endrin ketone

R.T.: 6.993 min
Delta R.T.: -0.002 min
Response: 830565711
Conc: 44.41 ng/ml




#22 Mirex

R.T.: 8.114 min
Delta R.T.: 0.000 min
Response: 118928969
Conc: 50.64 ng/ml



#22 Mirex


R.T.: 7.187 min
Delta R.T.: -0.003 min
Response: 642022830
Conc: 43.32 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.073 min
Delta R.T.: -0.002 min
Response: 160599924
Conc: 48.55 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050



#28 Decachlorobiphenyl

R.T.: 8.074 min
Delta R.T.: -0.003 min
Response: 768545575
Conc: 41.59 ng/ml