

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051521\
 Data File : PD062956.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 May 2021 18:30
 Operator : AR\AJ
 Sample : INDA339
 Misc :
 ALS Vial : 99 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA3039

Manual Integrations
 APPROVED

Ankita
 5/17/2021 5:12:43 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 15 01:19:47 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD042921CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Apr 30 02:49:51 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.441	3.971	20585423	237.5E6	21.501	19.561
27)	SA Decachlor...	8.198	9.027	48047136	475.4E6	44.762	43.551
Target Compounds							
2)	A alpha-BHC	3.877	4.359	27134179	333.8E6	19.380	17.649
3)	MA gamma-BHC...	4.159	4.644	26832268	306.4E6	20.087m	17.735
4)	MA Heptachlor	4.461	5.151	25397256	314.6E6	19.779	18.075
9)	A Endosulfan I	5.490	6.198	21721825	269.1E6	19.426m	18.345
13)	MA Dieldrin	5.734	6.455	48957033	568.8E6	38.839	36.681
14)	MA Endrin	5.992	6.673	42799486	471.9E6	40.450	37.841
16)	A 4,4'-DDD	6.118	6.795	38841304	453.9E6	39.006	36.521
17)	MA 4,4'-DDT	6.359	7.094	38121785	450.5E6	40.700	37.938
20)	A Methoxychlor	6.907	7.552	112.9E6	1078.3E6	211.527	190.230

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051521\
Data File : PD062956.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 May 2021 18:30
Operator : AR\AJ
Sample : INDA339
Misc :
ALS Vial : 99 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 15 01:19:47 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD042921CLP.M
Quant Title : GC Extractables
QLast Update : Fri Apr 30 02:49:51 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Manual Integrations
APPROVED

Ankita
5/17/2021 5:12:43 PM

Instrument :
ECD_D
ClientSampled :
INDA3039