

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051522\
 Data File : PD069632.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 May 2022 11:45
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 13 23:20:57 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041622.M
 Quant Title : GC Extractables
 QLast Update : Mon Apr 18 13:40:55 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.443	4.170	57750953	527.1E6	52.024	49.166
28)	SA Decachlor...	8.195	9.363	61693857	338.9E6	49.409	51.785
Target Compounds							
2)	A alpha-BHC	3.882	4.566	89103461	792.4E6	54.605	51.388
3)	MA gamma-BHC...	4.165	4.857	83227257	706.2E6	48.654	50.015
4)	MA Heptachlor	4.462	5.383	79517322	676.7E6	53.619	49.531
5)	MB Aldrin	4.709	5.696	80322574	661.7E6	55.156	50.075
6)	B beta-BHC	4.415	5.020	36112138	307.5E6	51.943	50.050
7)	B delta-BHC	4.619	5.245	81133590	648.6E6	54.694	48.363
8)	B Heptachlo...	5.153	6.082	77300600	626.9E6	56.392	52.975m
9)	A Endosulfan I	5.491	6.445	68773312	544.7E6	53.293	50.431m
10)	B gamma-Chl...	5.377	6.319	73971437	587.6E6	53.214m	50.466m
11)	B alpha-Chl...	5.438	6.395	73502948	619.3E6	53.446	54.149
12)	B 4,4'-DDE	5.603	6.545	70700306	556.9E6	55.875	51.550
13)	MA Dieldrin	5.735	6.705	71758351	568.8E6	53.180	51.026
14)	MA Endrin	5.993	6.924	63009079	376.3E6	52.114	41.489
15)	B Endosulfa...	6.267	7.130	59994908	466.5E6	51.845	53.917
16)	A 4,4'-DDD	6.120	7.036	59809424	470.5E6	54.873	54.630
17)	MA 4,4'-DDT	6.360	7.339	54952751	445.9E6	51.111	51.412
18)	B Endrin al...	6.436	7.254	47699298	358.5E6	51.679	51.855
19)	B Endosulfa...	6.649	7.479	62905820	442.6E6	52.005	51.847
20)	A Methoxychlor	6.910	7.792	31811525	214.4E6	50.077	49.728
21)	B Endrin ke...	7.142	7.954	73168888	472.5E6	51.967	52.260
22)	Mirex	7.324	8.420	56851574	327.8E6	49.974	50.856

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051522\
Data File : PD069632.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 May 2022 11:45
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 13 23:20:57 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041622.M
Quant Title : GC Extractables
QLast Update : Mon Apr 18 13:40:55 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

