

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051523\
 Data File : PD075461.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 May 2023 10:51
 Operator : AR\AJ
 Sample : 02592-11MS
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 JREF6MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 15 21:52:22 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu May 04 02:14:07 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.548	2.783	29161216	13204305	12.785	12.519
2)	SA Decachlor...	9.090	7.930	44089990	19107954	17.816	20.233
Target Compounds							
2)	A alpha-BHC	4.004	3.286	150.0E6	67190593	39.011	43.377
3)	MA gamma-BHC...	4.338	3.615	135.8E6	60203953	37.761	40.957
4)	MA Heptachlor	4.930	3.956	142.5E6	58164508	38.214	39.806
5)	MB Aldrin	5.273	4.236	149.9E6	63041553	41.094	42.507
6)	B beta-BHC	4.534	3.912	58146185	26852672	35.423	36.878
7)	B delta-BHC	4.783	4.143	109.5E6	49335048	32.381	31.223
8)	B Heptachlo...	5.700	4.739	127.3E6	53880380	37.566	38.630
9)	A Endosulfan I	6.087	5.109	118.1E6	49439642	39.150	41.194
10)	B trans-Chl...	5.957	4.990	125.3E6	54607813	38.244	40.019
11)	B cis-Chlor...	6.036	5.053	129.3E6	54937597	37.947	39.365
12)	B 4,4'-DDE	6.208	5.243	254.9E6	104.3E6	78.045	81.122
13)	MA Dieldrin	6.362	5.374	276.7E6	116.2E6	80.899	85.159
14)	MA Endrin	6.593	5.650	224.9E6	99022322	77.774	81.411
15)	B Endosulfa...	6.812	5.944	225.0E6	96202739	69.860	75.189
16)	A 4,4'-DDD	6.726	5.798	203.0E6	84762549	84.009	85.500
17)	MA 4,4'-DDT	7.042	6.049	209.1E6	81106315	82.396	80.473
18)	B Endrin al...	6.942	6.123	152.1E6	70635740	64.302	69.559
19)	B Endosulfa...	7.178	6.347	198.0E6	86566423	71.988	74.432
20)	A Methoxychlor	7.517	6.624	570.0E6	205.9E6	388.064	359.442
21)	B Endrin ke...	7.664	6.853	235.8E6	104.5E6	77.788	79.280

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051523\
Data File : PD075461.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 May 2023 10:51
Operator : AR\AJ
Sample : 02592-11MS
Misc :
ALS Vial : 13 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 15 21:52:22 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

