

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051523\
 Data File : PD075482.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 May 2023 15:59
 Operator : AR\AJ
 Sample : 02612-02MS
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 JREG4MS

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 05/16/2023
 Supervised By : Ankita Jodhani 05/16/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 15 22:09:13 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu May 04 02:14:07 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.548	2.770	24393498	11738794	10.695	11.130
27)	SA Decachlor...	9.089	7.915	34487178	15060193	13.936	15.947
Target Compounds							
2)	A alpha-BHC	4.003	3.272	143.9E6	64230196	37.430m	41.466
3)	MA gamma-BHC...	4.338	3.602	131.0E6	57379023	36.449	39.035
4)	MA Heptachlor	4.929	3.943	136.3E6	57150659	36.574	39.112
5)	MB Aldrin	5.272	4.222	143.4E6	58262298	39.310	39.284m
6)	B beta-BHC	4.533	3.899	56396854	25925651	34.357	35.605
7)	B delta-BHC	4.782	4.130	104.4E6	46634818	30.879	29.514
8)	B Heptachlo...	5.699	4.726	122.3E6	52548008	36.075	37.675
9)	A Endosulfan I	6.086	5.096	112.4E6	48108383	37.277	40.085
10)	B trans-Chl...	5.956	4.976	120.3E6	52109949	36.720	38.188
11)	B cis-Chlor...	6.035	5.040	123.0E6	52893611	36.083	37.901
12)	B 4,4'-DDE	6.207	5.230	241.8E6	100.5E6	74.056	78.187
13)	MA Dieldrin	6.362	5.361	262.4E6	110.7E6	76.742	81.159
14)	MA Endrin	6.593	5.636	211.9E6	94066834	73.283	77.337
15)	B Endosulfa...	6.811	5.930	226.3E6	92045116	70.269	71.940
16)	A 4,4'-DDD	6.725	5.784	195.7E6	83100156	80.964	83.823
17)	MA 4,4'-DDT	7.040	6.036	198.6E6	79971752	78.257	79.347
18)	B Endrin al...	6.941	6.110	152.4E6	67393916	64.421	66.367
19)	B Endosulfa...	7.177	6.333	187.9E6	82929597	68.308	71.305
20)	A Methoxychlor	7.516	6.612	533.0E6	208.9E6	362.840	364.644
21)	B Endrin ke...	7.663	6.840	225.7E6	101.9E6	74.464	77.290

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051523\
Data File : PD075482.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 May 2023 15:59
Operator : AR\AJ
Sample : 02612-02MS
Misc :
ALS Vial : 33 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 15 22:09:13 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Instrument :
ECD_D
ClientSampled :
JREG4MS

Manual Integrations
APPROVED

Reviewed By :Abdul Miza 05/16/2023
Supervised By :Ankita Jodhani 05/16/2023