

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051523\
Data File : PD075475.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 May 2023 14:04
Operator : AR\AJ
Sample : 02592-26
Misc :
ALS Vial : 27 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
JREK8

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/16/2023
Supervised By :Ankita Jodhani 05/16/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 15 22:05:38 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

3.548min 9.915 ng/ml

response 22614558

(1) Tetrachloro-m-xylene #2 (SA)

2.770min 15.879 ng/ml

response 16748332

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051523\
Data File : PD075475.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 May 2023 14:04
Operator : AR\AJ
Sample : 02592-26
Misc :
ALS Vial : 27 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
JREK8

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/16/2023
Supervised By :Ankita Jodhani 05/16/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 15 22:05:38 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

3.546min 11.255 ng/ml m

response 25672032

(1) Tetrachloro-m-xylene #2 (SA)

2.770min 15.879 ng/ml

response 16748332

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051523\
Data File : PD075475.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 May 2023 14:04
Operator : AR\AJ
Sample : 02592-26
Misc :
ALS Vial : 27 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
JREK8

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/16/2023
Supervised By :Ankita Jodhani 05/16/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 15 22:05:38 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(12) 4,4'-DDE (B)

6.205min 2.536 ng/ml

response 8281333

(12) 4,4'-DDE #2 (B)

5.231min 0.319 ng/ml

response 409685

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051523\
Data File : PD075475.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 May 2023 14:04
Operator : AR\AJ
Sample : 02592-26
Misc :
ALS Vial : 27 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
JREK8

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/16/2023
Supervised By :Ankita Jodhani 05/16/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 15 22:05:38 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(12) 4,4'-DDE (B)
6.205min 2.536 ng/ml
response 8281333

(12) 4,4'-DDE #2 (B)
5.229min 0.971 ng/ml m
response 1248589

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051523\
Data File : PD075475.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 May 2023 14:04
Operator : AR\AJ
Sample : 02592-26
Misc :
ALS Vial : 27 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
JREK8

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/16/2023
Supervised By :Ankita Jodhani 05/16/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 15 22:05:38 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(16) 4,4'-DDD (A)

6.734min 4.526 ng/ml

response 10937836

(16) 4,4'-DDD #2 (A)

5.784min 1.658 ng/ml

response 1643731

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051523\
Data File : PD075475.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 May 2023 14:04
Operator : AR\AJ
Sample : 02592-26
Misc :
ALS Vial : 27 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
JREK8

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/16/2023
Supervised By :Ankita Jodhani 05/16/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 15 22:05:38 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(16) 4,4'-DDD (A)

6.733min 4.058 ng/ml m
response 9806466

(16) 4,4'-DDD #2 (A)

5.784min 1.658 ng/ml
response 1643731

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051523\
Data File : PD075475.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 May 2023 14:04
Operator : AR\AJ
Sample : 02592-26
Misc :
ALS Vial : 27 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
JREK8

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/16/2023
Supervised By :Ankita Jodhani 05/16/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 15 22:05:38 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(17) 4,4'-DDT (MA)

7.045min 0.578 ng/ml

response 1468205

(17) 4,4'-DDT #2 (MA)

6.030min 9.939 ng/ml

response 10016945

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051523\
Data File : PD075475.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 May 2023 14:04
Operator : AR\AJ
Sample : 02592-26
Misc :
ALS Vial : 27 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
JREK8

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/16/2023
Supervised By :Ankita Jodhani 05/16/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 15 22:05:38 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(17) 4,4'-DDT (MA)

7.012min 7.939 ng/ml m

response 20151832

(17) 4,4'-DDT #2 (MA)

6.029min 8.946 ng/ml m

response 9016546

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051523\
Data File : PD075475.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 May 2023 14:04
Operator : AR\AJ
Sample : 02592-26
Misc :
ALS Vial : 27 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
JREK8

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/16/2023
Supervised By :Ankita Jodhani 05/16/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 15 22:05:38 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(27) Decachlorobiphenyl (SA)

9.089min 17.472 ng/ml

response 43237951

(27) Decachlorobiphenyl #2 (SA)

7.915min 21.962 ng/ml

response 20740571

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051523\
Data File : PD075475.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 May 2023 14:04
Operator : AR\AJ
Sample : 02592-26
Misc :
ALS Vial : 27 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
JREK8

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/16/2023
Supervised By :Ankita Jodhani 05/16/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 15 22:05:38 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(27) Decachlorobiphenyl (SA)

9.088min 18.487 ng/ml m

response 45750346

(27) Decachlorobiphenyl #2 (SA)

7.914min 25.027 ng/ml m

response 23635538