

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD051619CLP.M
 Title : GC Extractables
 Last Update : Fri May 17 11:03:48 2019
 Response Via : Initial Calibration

Calibration Files

3 =PD053279.D 1 =PD053277.D 2 =PD053278.D
 4 =PD053280.D 5 =PD053281.D

Compound		3	1	2	4	5	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	7.369	7.842	7.350	6.929	7.030	7.304	E5	4.90
2) A	alpha-BHC						0.000		-1.00
3) MA	gamma-BHC (Lindane)						0.000		-1.00
4) MA	Heptachlor						0.000		-1.00
5) MB	Aldrin						0.000		-1.00
6) B	beta-BHC						0.000		-1.00
7) B	delta-BHC						0.000		-1.00
8) B	Heptachlor epoxide						0.000		-1.00
9) A	Endosulfan I						0.000		-1.00
10) B	trans-Chlordane						0.000		-1.00
11) B	cis-Chlordane						0.000		-1.00
12) B	4,4'-DDE						0.000		-1.00
13) MA	Dieldrin						0.000		-1.00
14) MA	Endrin						0.000		-1.00
15) B	Endosulfan II						0.000		-1.00
16) A	4,4'-DDD						0.000		-1.00
17) MA	4,4'-DDT						0.000		-1.00
18) B	Endrin aldehyde						0.000		-1.00
19) B	Endosulfan Sulfate						0.000		-1.00
20) A	Methoxychlor						0.000		-1.00
21) B	Endrin ketone						0.000		-1.00
22)	Toxaphene-1	4.763	5.537	4.883	4.551	4.532	4.853	E3	8.44
23)	Toxaphene-2	4.446	4.994	4.551	4.196	4.187	4.475	E3	7.39
24)	Toxaphene-3	4.424	5.181	4.778	4.358	4.056	4.559	E3	9.48
25)	Toxaphene-4	1.073	1.161	1.123	0.997	1.054	1.082	E4	5.83
26)	Toxaphene-5	1.107	1.244	1.131	0.972	1.040	1.099	E4	9.30
27) SA	Decachlorobiphenyl	5.926	6.620	6.345	5.588	5.272	5.950	E5	9.19

Signal #2 Calibration Files

3 =PD053279.D 1 =PD053277.D 2 =PD053278.D
 4 =PD053280.D 5 =PD053281.D

Compound		3	1	2	4	5	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	0.908	1.015	0.931	0.833	0.817	0.901	E6	8.88
2) A	alpha-BHC						0.000		-1.00
3) MA	gamma-BHC (Lindane)						0.000		-1.00
4) MA	Heptachlor						0.000		-1.00
5) MB	Aldrin						0.000		-1.00
6) B	beta-BHC						0.000		-1.00
7) B	delta-BHC						0.000		-1.00
8) B	Heptachlor epoxide						0.000		-1.00
9) A	Endosulfan I						0.000		-1.00
10) B	trans-Chlordane						0.000		-1.00
11) B	cis-Chlordane						0.000		-1.00
12) B	4,4'-DDE						0.000		-1.00
13) MA	Dieldrin						0.000		-1.00
14) MA	Endrin						0.000		-1.00
15) B	Endosulfan II						0.000		-1.00
16) A	4,4'-DDD						0.000		-1.00
17) MA	4,4'-DDT						0.000		-1.00
18) B	Endrin aldehyde						0.000		-1.00
19) B	Endosulfan Sulfate						0.000		-1.00
20) A	Methoxychlor						0.000		-1.00
21) B	Endrin ketone						0.000		-1.00
22)	Toxaphene-1	4.075	4.433	4.209	3.672	3.919	4.062	E3	7.09
23)	Toxaphene-2	4.771	5.172	4.745	4.593	4.922	4.840	E3	4.53

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD051619CLP.M
 Title : GC Extractables
 Last Update : Fri May 17 11:03:48 2019
 Response Via : Initial Calibration

Calibration Files

3 =PD053279.D 1 =PD053277.D 2 =PD053278.D
 4 =PD053280.D 5 =PD053281.D

	Compound	3	1	2	4	5	Avg		%RSD
24)	Toxaphene-3	6.387	6.778	6.735	6.020	6.499	6.484	E3	4.72
25)	Toxaphene-4	1.785	1.876	1.814	1.743	1.909	1.825	E4	3.68
26)	Toxaphene-5	8.008	8.944	8.329	7.819	8.451	8.310	E3	5.23
27) SA	Decachlorobiphenyl	6.803	7.667	7.719	6.480	7.011	7.136	E5	7.60

(#) = Out of Range