

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051622\
 Data File : PD069649.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 May 2022 13:22
 Operator : AR\AJ
 Sample : N2806-02MSD
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 WC-8/9-1MSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 17 00:52:05 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041622.M
 Quant Title : GC Extractables
 QLast Update : Mon Apr 18 13:40:55 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.445	4.173	21926039	208.9E6	19.752	19.481
28)	SA Decachlor...	8.200	9.370	20652908	121.5E6	16.540	18.566
Target Compounds							
2)	A alpha-BHC	3.884	4.569	99838823	834.0E6	61.183	54.086
3)	MA gamma-BHC...	4.167	4.860	96057905	765.1E6	56.155	54.180
4)	MA Heptachlor	4.465	5.387	87499223	739.2E6	59.001	54.108
5)	MB Aldrin	4.712	5.700	88678447	744.3E6	60.894	56.322
6)	B beta-BHC	4.418	5.024	37462426	314.8E6	53.885	51.246
7)	B delta-BHC	4.622	5.250	88220496	704.0E6	59.471	52.493
8)	B Heptachlo...	5.156	6.088	81247247	638.8E6	59.271	53.982
9)	A Endosulfan I	5.494	6.451	74556387	593.3E6	57.774	54.931
10)	B gamma-Chl...	5.382	6.325	81630767	656.6E6	58.724	56.392
11)	B alpha-Chl...	5.440	6.400	80212089	635.7E6	58.324	55.577
12)	B 4,4'-DDE	5.606	6.550	75668764	603.2E6	59.802	55.840
13)	MA Dieldrin	5.738	6.710	82958263	648.4E6	61.481	58.161
14)	MA Endrin	5.996	6.930	75030225	561.1E6	62.056	61.867
15)	B Endosulfa...	6.270	7.136	71823665	549.1E6	62.067	63.463
16)	A 4,4'-DDD	6.123	7.041	64917103	514.4E6	59.559	59.724
17)	MA 4,4'-DDT	6.363	7.344	60924812	480.9E6	56.665	55.449
18)	B Endrin al...	6.439	7.259	57304629	430.9E6	62.086	62.330
19)	B Endosulfa...	6.652	7.484	70046577	494.8E6	57.908	57.960
20)	A Methoxychlor	6.914	7.797	34811332	245.4E6	54.799	56.928
21)	B Endrin ke...	7.146	7.960	80646900	520.7E6	57.278	57.595
22)	Mirex	7.329	8.425	57381309	345.7E6	50.440	53.624

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051622\
Data File : PD069649.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 May 2022 13:22
Operator : AR\AJ
Sample : N2806-02MSD
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
WC-8/9-1MSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 17 00:52:05 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041622.M
Quant Title : GC Extractables
QLast Update : Mon Apr 18 13:40:55 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

