

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051721\
 Data File : PD063017.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 May 2021 18:16
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 5/20/2021 1:33:59 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 18 08:10:37 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051721.M
 Quant Title : GC Extractables
 QLast Update : Tue May 18 07:12:39 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.440	3.970	55817019	573.0E6	50.673	49.450
28)	SA Decachlor...	8.196	9.024	65120208	562.9E6	51.742	55.171
Target Compounds							
2)	A alpha-BHC	3.876	4.358	89987702	957.4E6	51.900	49.725
3)	MA gamma-BHC...	4.159	4.642	83481206	863.6E6	51.345	50.036
4)	MA Heptachlor	4.460	5.150	80782172	847.0E6	51.353	51.082
5)	MB Aldrin	4.708	5.454	78913595	835.9E6	51.511	50.080
6)	B beta-BHC	4.410	4.813	38910279	371.3E6	49.986	50.054
7)	B delta-BHC	4.614	5.028	85627644	734.1E6	51.946	52.543m
8)	B Heptachlo...	5.151	5.840	73552893	746.8E6	51.039	48.828
9)	A Endosulfan I	5.490	6.196	69104888	721.7E6	50.925	50.804
10)	B gamma-Chl...	5.378	6.074	74973092	791.3E6	51.113	50.759
11)	B alpha-Chl...	5.437	6.147	73744044	752.7E6	51.078	50.794
12)	B 4,4'-DDE	5.600	6.303	73666247	769.3E6	51.349	50.964
13)	MA Dieldrin	5.733	6.453	75996178	773.8E6	51.362	51.032
14)	MA Endrin	5.990	6.671	65090603	633.2E6	48.667	49.628
15)	B Endosulfa...	6.264	6.879	63568115	626.8E6	51.095	52.266
16)	A 4,4'-DDD	6.116	6.793	62850548	634.8E6	52.246	52.319
17)	MA 4,4'-DDT	6.357	7.092	60817410	609.3E6	53.210	54.038
18)	B Endrin al...	6.432	7.004	53494680	526.4E6	50.326	51.689
19)	B Endosulfa...	6.646	7.227	63067705	650.8E6	52.326	53.039
20)	A Methoxychlor	6.905	7.550	34374875	297.2E6	51.224	51.814
21)	B Endrin ke...	7.139	7.697	80177830	649.6E6	54.258	54.156
22)	Mirex	7.326	8.151	55304676	399.4E6	51.542	52.516

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051721\
Data File : PD063017.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 May 2021 18:16
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
5/20/2021 1:33:59 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 18 08:10:37 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051721.M
Quant Title : GC Extractables
QLast Update : Tue May 18 07:12:39 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

