

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051925\
 Data File : PD088586.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 11:31
 Operator : AR\AJ
 Sample : PSTDICC100
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC100

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 12:44:12 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051925.M
 Quant Title : GC Extractables
 QLast Update : Mon May 19 12:41:43 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.550	2.881	214.9E6	1410.5E6	100.640	98.439
2)	SA Decachlor...	9.076	8.076	317.2E6	1650.4E6	97.585	97.596
Target Compounds							
2)	A alpha-BHC	4.000	3.394	524.4E6	2281.4E6	103.930	99.026
3)	MA gamma-BHC...	4.331	3.731	495.1E6	2103.5E6	103.492	99.023
4)	MA Heptachlor	4.931	4.085	468.1E6	2076.3E6	102.584	97.913
5)	MB Aldrin	5.273	4.371	459.6E6	2042.9E6	102.512	98.179
6)	B beta-BHC	4.516	4.027	176.6E6	886.5E6	100.163	97.589
7)	B delta-BHC	4.765	4.264	458.1E6	2112.6E6	104.014	98.953
8)	B Heptachlo...	5.693	4.875	401.8E6	1812.7E6	101.536	97.446
9)	A Endosulfan I	6.077	5.250	380.1E6	1729.7E6	101.233	97.439
10)	B gamma-Chl...	5.948	5.128	412.4E6	1982.5E6	102.141	98.414
11)	B alpha-Chl...	6.029	5.193	407.7E6	1904.9E6	101.529	98.142
12)	B 4,4'-DDE	6.198	5.378	377.5E6	1930.9E6	102.541	98.124
13)	MA Dieldrin	6.349	5.515	416.9E6	1938.3E6	102.216	97.907
14)	MA Endrin	6.576	5.792	354.4E6	1755.8E6	102.198	97.496
15)	B Endosulfa...	6.788	6.083	342.5E6	1667.1E6	100.933	97.341
16)	A 4,4'-DDD	6.707	5.932	291.4E6	1606.4E6	102.370	97.977
17)	MA 4,4'-DDT	7.023	6.187	325.9E6	1728.0E6	102.124	98.564
18)	B Endrin al...	6.917	6.261	250.3E6	1243.2E6	99.832	96.806
19)	B Endosulfa...	7.152	6.485	318.6E6	1609.8E6	100.650	97.278
20)	A Methoxychlor	7.495	6.757	160.6E6	857.2E6	98.864	96.006
21)	B Endrin ke...	7.632	6.994	343.4E6	1746.9E6	100.663	96.850
22)	Mirex	8.117	7.189	240.8E6	1351.3E6	97.999	96.782

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051925\
Data File : PD088586.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 11:31
Operator : AR\AJ
Sample : PSTDICC100
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 12:44:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051925.M
Quant Title : GC Extractables
QLast Update : Mon May 19 12:41:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

