

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052021\
 Data File : PD063126.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2021 15:19
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 07:15:09 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051721.M
 Quant Title : GC Extractables
 QLast Update : Tue May 18 07:12:39 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.445	3.971	56636370	552.5E6	51.417	47.681
28)	SA Decachlor...	8.200	9.026	56418224	499.3E6	44.828	48.946
Target Compounds							
2)	A alpha-BHC	3.881	4.359	91206049	919.3E6	52.603	47.744
3)	MA gamma-BHC...	4.164	4.643	84432904	826.2E6	51.931	47.870
4)	MA Heptachlor	4.465	5.151	79889317	794.8E6	50.785	47.935
5)	MB Aldrin	4.713	5.455	78863069	800.3E6	51.478	47.948
6)	B beta-BHC	4.415	4.814	38820334	406.0E6	49.850	54.729
7)	B delta-BHC	4.619	5.031	84731837	804.7E6	51.403	57.592
8)	B Heptachlo...	5.156	5.841	73289746	766.1E6	50.857	50.084
9)	A Endosulfan I	5.495	6.198	67784159	685.1E6	49.952	48.230
10)	B gamma-Chl...	5.383	6.075	74489158	756.0E6	50.784	48.498
11)	B alpha-Chl...	5.441	6.148	72474450	707.2E6	50.199	47.723
12)	B 4,4'-DDE	5.605	6.305	71335293	712.0E6	49.724	47.173
13)	MA Dieldrin	5.738	6.454	73609990	715.4E6	49.749	47.180
14)	MA Endrin	5.996	6.672	63788000	575.1E6	47.693	45.074
15)	B Endosulfa...	6.269	6.881	60752747	554.8E6	48.832	46.264
16)	A 4,4'-DDD	6.121	6.795	61834986	589.1E6	51.402	48.554
17)	MA 4,4'-DDT	6.362	7.093	53536078	521.5E6	46.840	46.253
18)	B Endrin al...	6.437	7.005	50971906	475.5E6	47.953	46.689
19)	B Endosulfa...	6.651	7.229	60312269	575.8E6	50.040	46.927
20)	A Methoxychlor	6.909	7.551	29770850	250.1E6	44.364	43.591
21)	B Endrin ke...	7.143	7.698	76138360	582.8E6	51.524	48.585
22)	Mirex	7.330	8.152	51090314	366.6E6	47.614	48.207

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052021\
Data File : PD063126.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2021 15:19
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 07:15:09 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051721.M
Quant Title : GC Extractables
QLast Update : Tue May 18 07:12:39 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

