

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052220\
 Data File : PD058085.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 May 2020 13:34
 Operator : AJ\MA
 Sample : L2735-02
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 C5CB7

Manual Integrations
 APPROVED

mohammad
 5/26/2020 1:49:31 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 22 16:04:08 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051820CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue May 19 05:31:06 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.209	3.867	16435474	25853222	20.965	22.019
27)	SA Decachlor...	7.851	8.851	28916364	36679997	33.424	32.641
Target Compounds							
8)	B Heptachlo...	4.849	5.705	7256085	17009911	7.922m	12.705m#
10)	B trans-Chl...	5.070	5.939	1771307	4413152	1.896	3.210m#
11)	B cis-Chlor...	5.125	6.014	3576314	5066485	3.804	3.656
13)	MA Dieldrin	5.416	6.316	24145701	33302941	24.552	23.807
17)	MA 4,4'-DDT	6.057	6.949	12864903	4184997	18.130m	3.837m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052220\
Data File : PD058085.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 May 2020 13:34
Operator : AJ\MA
Sample : L2735-02
Misc :
ALS Vial : 12 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 22 16:04:08 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051820CLP.M
Quant Title : GC Extractables
QLast Update : Tue May 19 05:31:06 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Manual Integrations
APPROVED

mohammad
5/26/2020 1:49:31 PM

Instrument :
ECD_D
ClientSample :
C5CB7