

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052225\
 Data File : PD088686.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 May 2025 12:40
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 21 14:22:33 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051925.M
 Quant Title : GC Extractables
 QLast Update : Mon May 19 15:27:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.551	2.881	119.1E6	812.0E6	55.051	53.679
28)	SA Decachlor...	9.075	8.074	180.9E6	915.5E6	52.860	50.150
Target Compounds							
2)	A alpha-BHC	4.000	3.393	264.6E6	1253.2E6	55.399	52.427
3)	MA gamma-BHC...	4.331	3.730	251.5E6	1151.5E6	54.637	51.949
4)	MA Heptachlor	4.931	4.083	238.4E6	1123.1E6	53.322	50.016
5)	MB Aldrin	5.273	4.370	239.9E6	1131.0E6	54.601	51.574
6)	B beta-BHC	4.516	4.026	96234882	500.3E6	53.370	51.358
7)	B delta-BHC	4.765	4.263	244.9E6	1155.4E6	57.939	51.826
8)	B Heptachlo...	5.692	4.873	212.2E6	1013.1E6	53.458	50.982
9)	A Endosulfan I	6.075	5.247	201.0E6	967.4E6	53.413	50.973
10)	B gamma-Chl...	5.947	5.126	215.4E6	1088.2E6	53.993	51.014
11)	B alpha-Chl...	6.028	5.191	215.2E6	1047.2E6	53.716	50.774
12)	B 4,4'-DDE	6.197	5.376	192.3E6	1051.0E6	53.563	50.279
13)	MA Dieldrin	6.348	5.513	215.6E6	1071.6E6	53.744	50.924
14)	MA Endrin	6.576	5.790	174.9E6	939.2E6	51.227	48.703
15)	B Endosulfa...	6.787	6.082	181.8E6	915.3E6	52.797	49.958
16)	A 4,4'-DDD	6.706	5.931	157.7E6	923.2E6	56.619	53.080
17)	MA 4,4'-DDT	7.023	6.185	137.3E6	802.3E6	43.994	44.235
18)	B Endrin al...	6.916	6.260	128.9E6	670.0E6	50.076	48.105
19)	B Endosulfa...	7.151	6.484	166.9E6	877.9E6	52.115	49.441
20)	A Methoxychlor	7.494	6.756	75154433	429.0E6	44.988	44.802
21)	B Endrin ke...	7.631	6.992	179.9E6	976.7E6	52.551	50.484
22)	Mirex	8.116	7.186	131.0E6	739.2E6	50.428	48.625

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052225\
Data File : PD088686.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 May 2025 12:40
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 21 14:22:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051925.M
Quant Title : GC Extractables
QLast Update : Mon May 19 15:27:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

