

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052322\
 Data File : PD069755.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 May 2022 19:12
 Operator : AR\AJ
 Sample : N2873-03
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 23 03:20:01 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD052022.M
 Quant Title : GC Extractables
 QLast Update : Mon May 23 02:08:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.444	4.172	18303493	196.8E6	18.265	20.603
28)	SA Decachlor...	8.197	9.368	12436226	85665961	11.529	14.294
Target Compounds							
3)	MA gamma-BHC...	4.173	0.000	10648338	0	7.212	N.D. #
16)	A 4,4'-DDD	0.000	7.050	0	8847029	N.D.	1.092 #
17)	MA 4,4'-DDT	0.000	7.317f	0	9971340	N.D.	1.212 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052322\
Data File : PD069755.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 May 2022 19:12
Operator : AR\AJ
Sample : N2873-03
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 23 03:20:01 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD052022.M
Quant Title : GC Extractables
QLast Update : Mon May 23 02:08:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.444 min
Delta R.T.: 0.000 min
Response: 18303493
Conc: 18.27 ng/ml

Instrument :
ECD_D
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.172 min
Delta R.T.: 0.000 min
Response: 196843166
Conc: 20.60 ng/ml




#3 gamma-BHC (Lindane)

R.T.: 4.173 min
Delta R.T.: 0.008 min
Response: 10648338
Conc: 7.21 ng/ml



#3 gamma-BHC (Lindane)

R.T.: 0.000 min
Exp R.T. : 4.860 min
Response: 0
Conc: N.D.



#16 4,4'-DDD

R.T.: 0.000 min
Exp R.T.: 6.122 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :

#16 4,4'-DDD


R.T.: 7.050 min
Delta R.T.: 0.009 min
Response: 8847029
Conc: 1.09 ng/ml

#17 4,4'-DDT

R.T.: 0.000 min
Exp R.T.: 6.363 min
Response: 0
Conc: N.D.

#17 4,4'-DDT


R.T.: 7.317 min
Delta R.T.: -0.027 min
Response: 9971340
Conc: 1.21 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.197 min
Delta R.T.: -0.002 min
Response: 12436226
Conc: 11.53 ng/ml

Instrument :
ECD_D
ClientSampleId :



#28 Decachlorobiphenyl

R.T.: 9.368 min
Delta R.T.: -0.003 min
Response: 85665961
Conc: 14.29 ng/ml