

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052422\
 Data File : PD069799.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 May 2022 20:19
 Operator : AR\AJ
 Sample : N2896-04MSD 10X
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 P038-SS001-0006-01MSD

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 05/25/2022
 Supervised By :Ankita Jodhani 05/25/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 24 02:41:40 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD052022.M
 Quant Title : GC Extractables
 QLast Update : Mon May 23 02:08:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.442	4.169	2559119	24340555	2.554	2.548m
28)	SA Decachlor...	8.194	9.363	2213570	15061203	2.052	2.513
Target Compounds							
2)	A alpha-BHC	3.881	4.565	8709720	94794937	5.832	6.796m
3)	MA gamma-BHC...	4.163	4.856	12115090	86378839	8.206m	6.745m
4)	MA Heptachlor	4.461	5.384	8447030	83182885	6.031	6.697
5)	MB Aldrin	4.707	5.697	8359932	81011071	6.205	6.724
6)	B beta-BHC	4.415	5.019	4546189	36423263	6.980	6.487m
7)	B delta-BHC	4.618	5.247	8038116	42403650	5.755	4.701
8)	B Heptachlo...	5.152	6.085	7802309	38782136	6.162	4.346 #
9)	A Endosulfan I	5.490	6.446	7189147	63630185	6.061	6.420m
10)	B gamma-Chl...	5.377	6.320	7971749	69015732	6.232	6.448m
11)	B alpha-Chl...	5.436	6.394	7818744	70942443	6.179	6.760m
12)	B 4,4'-DDE	5.601	6.545	6825184	63834679	5.897	6.431m
13)	MA Dieldrin	5.734	6.705	7980011	67693075	6.455	6.612
14)	MA Endrin	5.992	6.926	7369752	54036359	6.633	6.648
15)	B Endosulfa...	6.266	7.131	7098051	60843046	6.786	7.397
16)	A 4,4'-DDD	6.119	7.037	5887724	54765534	5.914	6.759
17)	MA 4,4'-DDT	6.358	7.340	5640085	53216760	5.875	6.471
18)	B Endrin al...	6.435	7.255	6026576	45102033	7.008	7.069
19)	B Endosulfa...	6.647	7.479	7175332	56111419	6.523	7.024
20)	A Methoxychlor	6.907	7.794	3468363	26645985	6.017m	6.462
21)	B Endrin ke...	7.141	7.953	7206459	54682837	5.753	6.546m
22)	Mirex	7.323	8.420	6026030	34794193	6.020	5.925

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052422\
Data File : PD069799.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 May 2022 20:19
Operator : AR\AJ
Sample : N2896-04MSD 10X
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
P038-SS001-0006-01MSD

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 05/25/2022
Supervised By : Ankita Jodhani 05/25/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 24 02:41:40 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD052022.M
Quant Title : GC Extractables
QLast Update : Mon May 23 02:08:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

