

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052524\
Data File : PD083001.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 May 2024 12:11
Operator : AR\AJ
Sample : P2547-16DL 5X
Misc :
ALS Vial : 11 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
BH5X3DL

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/28/2024
Supervised By :Ankita Jodhani 05/28/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 23 22:24:10 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051924CLP.M
Quant Title : GC Extractables
QLast Update : Fri May 17 05:17:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(1) Tetrachloro-m-xylene (SA)

3.564min 0.767 ng/ml

response 1147861

(1) Tetrachloro-m-xylene #2 (SA)

2.775min 2.485 ng/ml

response 9664166

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052524\
Data File : PD083001.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 May 2024 12:11
Operator : AR\AJ
Sample : P2547-16DL 5X
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

BH5X3DL


Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/28/2024

Supervised By :Ankita Jodhani 05/28/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 23 22:24:10 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051924CLP.M
Quant Title : GC Extractables
QLast Update : Fri May 17 05:17:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

3.563min 1.045 ng/ml m

response 1562869

(1) Tetrachloro-m-xylene #2 (SA)

2.775min 2.485 ng/ml

response 9664166

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052524\
Data File : PD083001.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 May 2024 12:11
Operator : AR\AJ
Sample : P2547-16DL 5X
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

BH5X3DL



Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/28/2024

Supervised By :Ankita Jodhani 05/28/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 23 22:24:10 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051924CLP.M
Quant Title : GC Extractables
QLast Update : Fri May 17 05:17:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(4) Heptachlor (MA)

4.974min 0.945 ng/ml

response 2379260

(4) Heptachlor #2 (MA)

3.924min 0.506 ng/ml

response 2816110

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052524\
Data File : PD083001.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 May 2024 12:11
Operator : AR\AJ
Sample : P2547-16DL 5X
Misc :
ALS Vial : 11 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
BH5X3DL

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/28/2024
Supervised By :Ankita Jodhani 05/28/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 23 22:24:10 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051924CLP.M
Quant Title : GC Extractables
QLast Update : Fri May 17 05:17:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(4) Heptachlor (MA)

4.972min 1.447 ng/ml m

response 3643122

(4) Heptachlor #2 (MA)

3.923min 0.910 ng/ml m

response 5066134

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052524\
Data File : PD083001.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 May 2024 12:11
Operator : AR\AJ
Sample : P2547-16DL 5X
Misc :
ALS Vial : 11 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
BH5X3DL

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/28/2024
Supervised By :Ankita Jodhani 05/28/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 23 22:24:10 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051924CLP.M
Quant Title : GC Extractables
QLast Update : Fri May 17 05:17:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(9) Endosulfan I (A)

6.106min 2.601 ng/ml

response 5413367

(9) Endosulfan I #2 (A)

5.120min 24.414 ng/ml

response 110104654

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052524\
Data File : PD083001.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 May 2024 12:11
Operator : AR\AJ
Sample : P2547-16DL 5X
Misc :
ALS Vial : 11 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
BH5X3DL

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/28/2024
Supervised By :Ankita Jodhani 05/28/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 23 22:24:10 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051924CLP.M
Quant Title : GC Extractables
QLast Update : Fri May 17 05:17:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(9) Endosulfan I (A)
6.138min 11.704 ng/ml m
response 24353599

(9) Endosulfan I #2 (A)
5.120min 24.414 ng/ml
response 110104654

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052524\
Data File : PD083001.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 May 2024 12:11
Operator : AR\AJ
Sample : P2547-16DL 5X
Misc :
ALS Vial : 11 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
BH5X3DL

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/28/2024
Supervised By :Ankita Jodhani 05/28/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 23 22:24:10 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051924CLP.M
Quant Title : GC Extractables
QLast Update : Fri May 17 05:17:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(13) Dieldrin (MA)

6.402min 7.179 ng/ml

response 15778830

(13) Dieldrin #2 (MA)

5.363min 8.985 ng/ml

response 44899011

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052524\
Data File : PD083001.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 May 2024 12:11
Operator : AR\AJ
Sample : P2547-16DL 5X
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

BH5X3DL


Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/28/2024

Supervised By :Ankita Jodhani 05/28/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 23 22:24:10 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051924CLP.M
Quant Title : GC Extractables
QLast Update : Fri May 17 05:17:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(13) Dieldrin (MA)

6.401min 3.332 ng/ml m

response 7322453

(13) Dieldrin #2 (MA)

5.362min 13.526 ng/ml m

response 67588498

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052524\
Data File : PD083001.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 May 2024 12:11
Operator : AR\AJ
Sample : P2547-16DL 5X
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

BH5X3DL



Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/28/2024

Supervised By :Ankita Jodhani 05/28/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 23 22:24:10 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051924CLP.M
Quant Title : GC Extractables
QLast Update : Fri May 17 05:17:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(14) Endrin (MA)

0.000min 0.000 ng/ml

response 0

(14) Endrin #2 (MA)

5.641min 15.358 ng/ml

response 67168357

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052524\
Data File : PD083001.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 May 2024 12:11
Operator : AR\AJ
Sample : P2547-16DL 5X
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

BH5X3DL



Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/28/2024

Supervised By :Ankita Jodhani 05/28/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 23 22:24:10 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051924CLP.M
Quant Title : GC Extractables
QLast Update : Fri May 17 05:17:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(14) Endrin (MA)

6.647min 16.821 ng/ml m

response 29995048

(14) Endrin #2 (MA)

5.639min 16.014 ng/ml m

response 70040994

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052524\
Data File : PD083001.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 May 2024 12:11
Operator : AR\AJ
Sample : P2547-16DL 5X
Misc :
ALS Vial : 11 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
BH5X3DL

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/28/2024
Supervised By :Ankita Jodhani 05/28/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 23 22:24:10 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051924CLP.M
Quant Title : GC Extractables
QLast Update : Fri May 17 05:17:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(16) 4,4'-DDD (A)
6.735min 9.365 ng/ml
response 14523734

(16) 4,4'-DDD #2 (A)
5.779min 1.967 ng/ml
response 7400593

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052524\
Data File : PD083001.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 May 2024 12:11
Operator : AR\AJ
Sample : P2547-16DL 5X
Misc :
ALS Vial : 11 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
BH5X3DL

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/28/2024
Supervised By :Ankita Jodhani 05/28/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 23 22:24:10 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051924CLP.M
Quant Title : GC Extractables
QLast Update : Fri May 17 05:17:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(16) 4,4'-DDD (A)
6.733min 10.459 ng/ml m
response 16220883

(16) 4,4'-DDD #2 (A)
5.777min 4.408 ng/ml m
response 16587995

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052524\
Data File : PD083001.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 May 2024 12:11
Operator : AR\AJ
Sample : P2547-16DL 5X
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

BH5X3DL



Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/28/2024

Supervised By :Ankita Jodhani 05/28/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 23 22:24:10 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051924CLP.M
Quant Title : GC Extractables
QLast Update : Fri May 17 05:17:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(27) Decachlorobiphenyl (SA)

9.103min 4.695 ng/ml

response 9411039

(27) Decachlorobiphenyl #2 (SA)

7.901min 6.678 ng/ml

response 26421799

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052524\
Data File : PD083001.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 May 2024 12:11
Operator : AR\AJ
Sample : P2547-16DL 5X
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

BH5X3DL



Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/28/2024

Supervised By :Ankita Jodhani 05/28/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 23 22:24:10 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051924CLP.M
Quant Title : GC Extractables
QLast Update : Fri May 17 05:17:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(27) Decachlorobiphenyl (SA)

9.101min 8.985 ng/ml m

response 18010502

(27) Decachlorobiphenyl #2 (SA)

7.901min 6.678 ng/ml

response 26421799