

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052622\
 Data File : PD069846.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 May 2022 00:46
 Operator : AR\AJ
 Sample : PTOXICC250
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PTOXICC250

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 05/26/2022
 Supervised By : Ankita Jodhani 05/26/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 26 04:44:24 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX052622.M
 Quant Title : GC Extractables
 QLast Update : Thu May 26 04:44:19 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.442	4.169	26933070	237.4E6	24.062	25.398
7)	SA Decachlor...	8.193	9.361	28377692	141.4E6	25.384	25.633
Target Compounds							
2)	Toxaphene-1	5.694	6.187	2187976	6659096	252.411	258.477
3)	Toxaphene-2	5.911	6.790	4221297	25227480	276.860	249.616
4)	Toxaphene-3	6.297	7.470	5235737	33994956	244.324	250.092
5)	Toxaphene-4	6.543	7.611	5340429	13743837	245.302	248.098m
6)	Toxaphene-5	7.030	7.881	5650717	23526510	241.892	243.620m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052622\
Data File : PD069846.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 May 2022 00:46
Operator : AR\AJ
Sample : PTOXICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 26 04:44:24 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX052622.M
Quant Title : GC Extractables
QLast Update : Thu May 26 04:44:19 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Reviewed By: Abdul Mirza 05/26/2022
Supervised By: Ankita Jodhani 05/26/2022

Manual Integrations
APPROVED

Instrument :
ECD_D
Client Sampled :
PTOXICC250