Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052624\
Data File : PD083078.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 May 2024 01:41
Operator : AR\AJ
Sample : P2568-15
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
BH635

Manual Integrations APPROVED

Reviewed By :Abdul Mirza 05/28/2024
Supervised By :Ankita Jodhani 05/28/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 25 02:28:42 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051924CLP.M
Quant Title : GC Extractables
QLast Update : Fri May 17 05:17:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.564	2.774	12303300	44485605	8.225m	11.440m#
27)	SA Decachlor...	9.119	7.903	17052104	31061512	8.507m	7.851
	Target Compounds						
9)	A Endosulfan I	6.139	5.118	4354241	10422756	2.093m	2.311m
12)	B 4,4'-DDE	6.227	5.222	7070370	22401253	3.242	4.697m#
13)	MA Dieldrin	6.402	5.359	3659873	19477218	1.665m	3.898m#
16)	A 4,4'-DDD	6.732	5.777	2437168	5487622	1.571m	1.458m
17)	MA 4,4'-DDT	7.064	6.027	4432291	18046038	2.756m	4.700 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052624\
Data File : PD083078.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 May 2024 01:41
Operator : AR\AJ
Sample : P2568-15
Misc :
ALS Vial : 37 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 25 02:28:42 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051924CLP.M
Quant Title : GC Extractables
QLast Update : Fri May 17 05:17:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm





Manual Integrations APPROVED
Reviewed By :Abdul Mirza 05/28/2024
Supervised By :Ankita Jodhani 05/28/2024

ClientSampleId : BH635

Instrument : ECD_D