

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052722\
Data File : PD069883.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 May 2022 15:41
Operator : AR\AJ
Sample : INDA270
Misc :
ALS Vial : 14 Sample Multiplier: 1


Instrument :
ECD_D
LabSampleId :
INDA270

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/27/2022
Supervised By :Ankita Jodhani 05/27/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 27 03:28:13 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050722CLP.M
Quant Title : GC Extractables
QLast Update : Wed May 18 23:46:25 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(4) Heptachlor (MA)

4.461min 16.615 ng/ml

response 26880742

(4) Heptachlor #2 (MA)

5.383min 16.401 ng/ml

response 259516663

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052722\
 Data File : PD069883.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 May 2022 15:41
 Operator : AR\AJ
 Sample : INDA270
 Misc :
 ALS Vial : 14 Sample Multiplier: 1


Instrument :
 ECD_D
 LabSampleId :
 INDA270

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/27/2022
 Supervised By :Ankita Jodhani 05/27/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 27 03:28:13 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050722CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed May 18 23:46:25 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(4) Heptachlor (MA)
 4.459min 17.121 ng/ml m
 response 27699902

(4) Heptachlor #2 (MA)
 5.383min 16.401 ng/ml
 response 259516663

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052722\
Data File : PD069883.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 May 2022 15:41
Operator : AR\AJ
Sample : INDA270
Misc :
ALS Vial : 14 Sample Multiplier: 1


Instrument :
ECD_D
LabSampleId :
INDA270

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/27/2022
Supervised By :Ankita Jodhani 05/27/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 27 03:28:13 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050722CLP.M
Quant Title : GC Extractables
QLast Update : Wed May 18 23:46:25 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(14) Endrin (MA)

5.990min 32.402 ng/ml

response 42190189

(14) Endrin #2 (MA)

6.924min 30.280 ng/ml

response 306230870

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052722\
Data File : PD069883.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 May 2022 15:41
Operator : AR\AJ
Sample : INDA270
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :

ECD_D

LabSampleId :

INDA270


Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 05/27/2022

Supervised By :Ankita Jodhani 05/27/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 27 03:28:13 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050722CLP.M
Quant Title : GC Extractables
QLast Update : Wed May 18 23:46:25 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(14) Endrin (MA)

5.989min 32.778 ng/ml m

response 42678873

(14) Endrin #2 (MA)

6.924min 30.280 ng/ml

response 306230870