

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052724\
 Data File : PD083151.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 May 2024 15:43
 Operator : AR\AJ
 Sample : P2569-13
 Misc :
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 BH698

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 05/28/2024
 Supervised By : Ankita Jodhani 05/28/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 27 21:59:51 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051924CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri May 17 05:17:58 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.563	2.775	27006942	77174360	18.056m	19.846
27)	SA Decachlor...	9.113	7.901	57161985	128.9E6	28.518m	32.583
Target Compounds							
4)	MA Heptachlor	4.972	3.922	6379560	9618160	2.534m	1.727m#
9)	A Endosulfan I	6.139	5.117	48854711	106.5E6	23.478m	23.607m
12)	B 4,4'-DDE	6.227	5.223	82154233	275.5E6	37.672	57.765 #
13)	MA Dieldrin	6.403	5.361	29103796	110.6E6	13.242	22.143m#
14)	MA Endrin	6.643	5.636	22121324	33621745	12.405m	7.687m#
16)	A 4,4'-DDD	6.731	5.776	35658113	56742851	22.992m	15.080m#
17)	MA 4,4'-DDT	7.062	6.026	127.2E6	284.5E6	79.075	74.099

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052724\
Data File : PD083151.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 May 2024 15:43
Operator : AR\AJ
Sample : P2569-13
Misc :
ALS Vial : 50 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 27 21:59:51 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051924CLP.M
Quant Title : GC Extractables
QLast Update : Fri May 17 05:17:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Manual Integrations
APPROVED
Reviewed By: Abul Mirza 05/28/2024
Supervised By: Ankita Jodhani 05/28/2024

Instrument :
ECD_D
ClientSampled :
BH698