

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052725\
 Data File : PD088749.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 May 2025 17:24
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 28 08:37:41 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051925.M
 Quant Title : GC Extractables
 QLast Update : Mon May 19 15:27:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.549	2.880	122.0E6	823.6E6	56.368	54.449
28)	SA Decachlor...	9.069	8.071	169.7E6	801.0E6	49.587	43.876
Target Compounds							
2)	A alpha-BHC	3.998	3.393	278.7E6	1306.3E6	58.353	54.649
3)	MA gamma-BHC...	4.329	3.729	260.6E6	1189.7E6	56.621	53.672
4)	MA Heptachlor	4.928	4.083	240.1E6	1186.2E6	53.697	52.825
5)	MB Aldrin	5.270	4.368	234.6E6	1160.2E6	53.385	52.904
6)	B beta-BHC	4.514	4.025	96940299	498.2E6	53.761	51.144
7)	B delta-BHC	4.762	4.262	210.6E6	1180.5E6	49.826	52.951
8)	B Heptachlo...	5.690	4.872	203.0E6	1008.8E6	51.155	50.765
9)	A Endosulfan I	6.072	5.246	192.7E6	915.6E6	51.215	48.246
10)	B gamma-Chl...	5.944	5.125	206.6E6	1077.8E6	51.782	50.525
11)	B alpha-Chl...	6.025	5.189	205.3E6	1033.5E6	51.247	50.109
12)	B 4,4'-DDE	6.194	5.374	184.2E6	1067.1E6	51.312	51.050
13)	MA Dieldrin	6.346	5.511	209.8E6	1044.0E6	52.303	49.610
14)	MA Endrin	6.572	5.788	180.6E6	959.1E6	52.869	49.738
15)	B Endosulfa...	6.784	6.079	187.5E6	883.3E6	54.449	48.213
16)	A 4,4'-DDD	6.703	5.928	144.2E6	840.3E6	51.767	48.315
17)	MA 4,4'-DDT	7.019	6.182	163.2E6	887.1E6	52.259	48.912
18)	B Endrin al...	6.913	6.257	123.5E6	642.1E6	47.973	46.096
19)	B Endosulfa...	7.147	6.481	161.0E6	841.1E6	50.252	47.373
20)	A Methoxychlor	7.491	6.753	97784705	524.1E6	58.534	54.740
21)	B Endrin ke...	7.628	6.990	173.2E6	882.7E6	50.599	45.621
22)	Mirex	8.111	7.184	124.9E6	668.0E6	48.067	43.940

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD052725\
Data File : PD088749.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 May 2025 17:24
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 28 08:37:41 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051925.M
Quant Title : GC Extractables
QLast Update : Mon May 19 15:27:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

