

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD053019\
 Data File : PD053436.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 May 2019 13:10
 Operator : SM\AJ
 Sample : PIBLK28
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PIBLK28

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 31 01:14:01 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD052119CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed May 22 06:53:42 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.317	3.986	15927410	23990343	21.738	24.367
27)	SA Decachlor...	8.016	9.031	26924793	41053264	40.702	39.744
Target Compounds							
8)	B Heptachlo...	0.000	5.845	0	1053721	N.D.	0.898
16)	A 4,4'-DDD	0.000	6.749f	0	270537	N.D.	0.293


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD053019\
Data File : PD053436.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 May 2019 13:10
Operator : SM\AJ
Sample : PIBLK28
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PIBLK28

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 31 01:14:01 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD052119CLP.M
Quant Title : GC Extractables
QLast Update : Wed May 22 06:53:42 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.317 min
Delta R.T.: 0.000 min
Response: 15927410
Conc: 21.74 ng/ml

Instrument :
ECD_D
ClientSampleId :
PIBLK28




#1 Tetrachloro-m-xylene

R.T.: 3.986 min
Delta R.T.: 0.000 min
Response: 23990343
Conc: 24.37 ng/ml




#8 Heptachlor epoxide

R.T.: 0.000 min
Exp R.T. : 4.992 min
Response: 0
Conc: N.D.



#8 Heptachlor epoxide


R.T.: 5.845 min
Delta R.T.: -0.006 min
Response: 1053721
Conc: 0.90 ng/ml



#16 4,4'-DDD


R.T.: 0.000 min
Exp R.T.: 5.953 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
PIBLK28




#16 4,4'-DDD

R.T.: 6.749 min
Delta R.T.: -0.050 min
Response: 270537
Conc: 0.29 ng/ml



#27 Decachlorobiphenyl

R.T.: 8.016 min
Delta R.T.: 0.002 min
Response: 26924793
Conc: 40.70 ng/ml



#27 Decachlorobiphenyl

R.T.: 9.031 min
Delta R.T.: 0.001 min
Response: 41053264
Conc: 39.74 ng/ml