

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD053024\
 Data File : PD083261.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 May 2024 23:35
 Operator : AR\AJ
 Sample : INDA338
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA3138

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 31 00:40:53 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051924CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri May 17 05:17:58 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.562	2.773	28927912	91720832	19.340	23.587
27)	SA Decachlor...	9.116	7.900	71045458	168.3E6	35.445	42.548
Target Compounds							
2)	A alpha-BHC	4.020	3.272	49318769	145.3E6	19.331	23.378
3)	MA gamma-BHC...	4.354	3.600	47739604	135.3E6	19.346	23.386
4)	MA Heptachlor	4.946	3.939	49509766	128.8E6	19.662	23.132
9)	A Endosulfan I	6.105	5.088	40787712	106.3E6	19.601	23.581
13)	MA Dieldrin	6.380	5.352	86080939	231.4E6	39.166	46.307
14)	MA Endrin	6.611	5.628	66002062	195.2E6	37.013	44.638
16)	A 4,4'-DDD	6.744	5.774	61728661	179.5E6	39.802	47.713
17)	MA 4,4'-DDT	7.060	6.025	58693947	165.1E6	36.493	43.009
20)	A Methoxychlor	7.536	6.600	157.6E6	425.5E6	182.172	216.844

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD053024\
Data File : PD083261.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2024 23:35
Operator : AR\AJ
Sample : INDA338
Misc :
ALS Vial : 19 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 31 00:40:53 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051924CLP.M
Quant Title : GC Extractables
QLast Update : Fri May 17 05:17:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

