

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD053123\
 Data File : PD075640.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 May 2023 13:31
 Operator : AR\AJ
 Sample : 02970-07
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEST-GPC-BLANK-SPIKE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 01 02:21:48 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu May 04 02:14:07 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds

3) MA	gamma-BHC...	4.337	3.600	68469519	31140837	19.045	21.185
4) MA	Heptachlor	4.928	3.941	69222982	29487673	18.569	20.181
5) MB	Aldrin	5.271	4.221	69459233	32153955	19.043	21.680
13) MA	Dieldrin	6.361	5.359	126.4E6	56338501	36.963	41.292
14) MA	Endrin	6.592	5.634	108.1E6	48509699	37.384	39.882
17) MA	4,4'-DDT	7.040	6.033	83188803	35722408	32.774	35.443


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD053123\
 Data File : PD075640.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 May 2023 13:31
 Operator : AR\AJ
 Sample : 02970-07
 Misc :
 ALS Vial : 5 Sample Multiplier: 1


Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 01 02:21:48 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
 Quant Title : GC Extractables
 Qlast Update : Thu May 04 02:14:07 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Signal: PD075640.D\ECD1A.ch



Signal: PD075640.D\ECD2B.ch



Instrument :
 ECD_D
 ClientSampleId :
 PEST-GPC-BLANK-SPIKE