

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD053125\
 Data File : PD088782.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 May 2025 10:36
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 30 22:49:10 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051925.M
 Quant Title : GC Extractables
 QLast Update : Mon May 19 15:27:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.558	2.880	113.4E6	736.7E6	52.409	48.701
28)	SA Decachlor...	9.088	8.078	172.7E6	890.0E6	50.477	48.750
Target Compounds							
2)	A alpha-BHC	4.008	3.394	258.4E6	1189.9E6	54.096	49.777
3)	MA gamma-BHC...	4.339	3.730	244.6E6	1099.7E6	53.142	49.611
4)	MA Heptachlor	4.938	4.084	242.8E6	1071.5E6	54.304	47.718
5)	MB Aldrin	5.281	4.370	233.3E6	1074.3E6	53.091	48.987
6)	B beta-BHC	4.524	4.026	92881314	446.8E6	51.510	45.869
7)	B delta-BHC	4.773	4.263	239.6E6	1105.9E6	56.705	49.607
8)	B Heptachlo...	5.701	4.875	203.3E6	970.7E6	51.227	48.850
9)	A Endosulfan I	6.085	5.250	198.8E6	927.7E6	52.835	48.884
10)	B gamma-Chl...	5.956	5.128	211.6E6	1049.2E6	53.058	49.185
11)	B alpha-Chl...	6.037	5.193	212.0E6	1009.7E6	52.926	48.957
12)	B 4,4'-DDE	6.207	5.378	191.1E6	1019.3E6	53.237	48.764
13)	MA Dieldrin	6.358	5.516	212.3E6	1029.3E6	52.912	48.912
14)	MA Endrin	6.585	5.792	175.6E6	914.7E6	51.418	47.434
15)	B Endosulfa...	6.797	6.084	177.3E6	903.4E6	51.489	49.307
16)	A 4,4'-DDD	6.716	5.933	148.0E6	854.0E6	53.138	49.102
17)	MA 4,4'-DDT	7.032	6.187	160.9E6	897.6E6	51.533	49.487
18)	B Endrin al...	6.926	6.262	131.6E6	670.5E6	51.123	48.141
19)	B Endosulfa...	7.160	6.486	164.5E6	861.3E6	51.360	48.506
20)	A Methoxychlor	7.504	6.759	83187231	443.8E6	49.796	46.350
21)	B Endrin ke...	7.642	6.995	179.1E6	975.2E6	52.309	50.406
22)	Mirex	8.126	7.189	130.2E6	728.3E6	50.103	47.905

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD053125\
Data File : PD088782.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2025 10:36
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 22:49:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051925.M
Quant Title : GC Extractables
QLast Update : Mon May 19 15:27:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

