

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060120\
 Data File : PD058228.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Jun 2020 09:52
 Operator : AJ\MA
 Sample : PSTDICC100
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC100

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 01 11:17:45 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060120.M
 Quant Title : GC Extractables
 QLast Update : Mon Jun 01 11:12:21 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.209	3.866	80004364	120.1E6	100.242	99.040
28)	SA Decachlor...	7.850	8.850	83128699	110.0E6	95.754	92.797
Target Compounds							
2)	A alpha-BHC	3.631	4.248	131.2E6	189.8E6	104.100	102.099
3)	MA gamma-BHC...	3.902	4.527	117.8E6	173.7E6	101.731	100.992
4)	MA Heptachlor	4.182	5.025	72355453	116.8E6	99.815	95.022
5)	MB Aldrin	4.417	5.326	103.0E6	158.8E6	101.226	100.676
6)	B beta-BHC	4.152	4.698	46674731	68566355	98.708	97.371
7)	B delta-BHC	4.343	4.909	116.0E6	165.1E6	102.228	101.645
8)	B Heptachlo...	4.851	5.709	82506890	139.3E6	99.184	98.138
9)	A Endosulfan I	5.178	6.061	84196763	133.7E6	99.302	98.219
10)	B gamma-Chl...	5.070	5.941	94966372	147.0E6	100.537	99.713
11)	B alpha-Chl...	5.127	6.013	96654512	146.3E6	100.410	99.168
12)	B 4,4'-DDE	5.297	6.173	101.9E6	146.1E6	102.083	101.011
13)	MA Dieldrin	5.415	6.315	56886923	124.6E6	95.726	96.504
14)	MA Endrin	5.666	6.531	50722447	106.0E6	94.477	94.161
15)	B Endosulfa...	5.941	6.740	65253440	121.3E6	96.205	96.093
16)	A 4,4'-DDD	5.804	6.658	40042067	90291118	95.446	94.811
17)	MA 4,4'-DDT	6.038	6.954	58789527	100.6E6	100.793	96.376
18)	B Endrin al...	6.108	6.863	42873100	96261620	94.332	93.017
19)	B Endosulfa...	6.319	7.085	62471634	107.8E6	97.600	95.927
20)	A Methoxychlor	6.587	7.413	19783861	43082471	93.426	90.382
21)	B Endrin ke...	6.799	7.550	53089578	111.5E6	91.441	91.105
22)	Mirex	6.970	8.001	67544849	95392510	97.120	93.328

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060120\
Data File : PD058228.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Jun 2020 09:52
Operator : AJ\MA
Sample : PSTDICC100
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 01 11:17:45 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060120.M
Quant Title : GC Extractables
QLast Update : Mon Jun 01 11:12:21 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

