

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060221\
 Data File : PD063365.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jun 2021 14:34
 Operator : AR\AJ
 Sample : PTOXICV500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 ICVPD060221

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 03 01:20:16 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX060221.M
 Quant Title : GC Extractables
 QLast Update : Thu Jun 03 01:19:15 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA Tetrachloro-m-xylene	1.374	1.355 E6	1.4	100	0.00
2 Toxaphene-1	10.721	10.714 E3	0.1	103	0.00
3 Toxaphene-2	11.584	12.293 E3	-6.1	106	0.00
4 Toxaphene-3	27.194	27.482 E3	-1.1	104	0.00
5 Toxaphene-4	28.722	28.583 E3	0.5	104	0.00
6 Toxaphene-5	31.397	31.748 E3	-1.1	101	0.00
7 SA Decachlorobiphenyl	1.604	1.594 E6	0.6	102	0.00

Signal #2

1 SA Tetrachloro-m-xylene	13.575	13.384 E6	1.4	101	0.00
2 Toxaphene-1	164.170	169.215 E3	-3.1	106	0.00
3 Toxaphene-2	138.402	145.703 E3	-5.3	108	0.00
4 Toxaphene-3	408.758	418.433 E3	-2.4	105	0.00
5 Toxaphene-4	251.254	258.591 E3	-2.9	103	0.00
6 Toxaphene-5	180.920	180.051 E3	0.5	102	0.00
7 SA Decachlorobiphenyl	14.218	14.166 E6	0.4	102	0.00

Evaluate Continuing Calibration Report - Not Found

Signal #2

(#) = Out of Range

SPCC's out = 0 CCC's out = 0